testo 815



Sound level measuring instrument

testo 815

Frequency evaluation according to characteristic curves A and C

Easy to adjust

Maximum and minimum store

Switchable time evaluation Fast/Slow





The testo 815 is the ideal instrument for the evaluation of sound level in everyday work. The instrument can be used for monitoring measurements in the areas of air conditioning or heating, on combustion plants or at workplaces, however machine or discotheque noise can also be evaluated quickly and precisely.

Using the Fast/slow button, the measurement time can be set from 1 s to 125 ms, in order to better evaluate the noise source. The frequency analysis can also be switched from characteristic curve A to C. The characteristic curve A

corresponds to the sound pressure perception of the human ear, whereas characteristic curve C can also evaluate the low-frequency components of a sound.

With the help of the optional calibrator, the measuring instrument can be recalibrated directly on site with the screwdriver provided. The wind protection cap included in delivery ensures correct measurement results when measuring outdoors, and if required can protect the microphone from dirt and dust.



99 Washington Street Melrose, MA 02176 Phone 781-665-1400 Toll Free 1-800-517-8431



Technical data / Accessories

testo 815

testo 815, sound level meter incl. calibration scewdriver, wind screen and battery

Part no. 0563 8155



Technical data testo 815

Meas. range	+32 to +130 dB
Frequency range	31.5 Hz to 8 kHz
Accuracy ±1 digit	±1.0 dB
Resolution	0.1 dB
Operating temperature	0 to +40 °C
Storage temperature	-10 to +60 °C
Battery type	9V block battery
Battery life	70 h
Weight	195 g
Dimensions	255 x 55 x 43 mm
Warranty	2 years

Section meas. ranges: 30 to 80 dB; 50 to 100 dB; 80 to 130 dB Time weighting: FAST 125 ms setting / SLOW 1 s setting Pressure dependency: -0.0016 dB/hPa

Technical data Sound level calibrator

Battery type	9V block battery
Battery life	40 h
Warranty	2 years
Accuracy	±0.5 dB in accordance with Class 2 to IEC 60942

Sound pressure level: 94 dB/104 dB, switchable

Frequency: 1000 Hz

Also suitable for 1/2 and 1 inch microphones by other manufacturers

Accessories Part no.

Accessories for measuring instrument

Calibrator, for regular calibration of testo 815, testo 816		
ISO calibration cert./sound pressure; calibration point 94 dB at different frequencies	0520 0111	
ISO calibration certificate sound pressure calibrators	0520 0411	

Test Equipment Depot - 800.517.8431 99 Washington Street, Melrose, MA 02176 TestEquipmentDepot.com