

BATTERY HITESTER BT3561A, BT3562A, BT3563A

5 Commonwealth Ave Woburn, MA 01801 Phone 781-665-1400 Toll Free 1-800-517-8431

Visit us at www.TestEquipmentDepot.com

Renewal of HIOKI's world-leading battery tester







Designing automatic battery testing systems is easier and faster than ever before

- Double the total line resistance, so measurement errors are less likely to occur when using long measurement cables
- Stable operation regardless of increased total line resistance due to probe and relay degradation
- LAN is equipped as a standard for easy system design and layout, and excellent noise resistance for stable operation
- Improved electrostatic resistance as a countermeasure against electrostatic charges during battery transport on a production line



Lineup

| Application Model Appearance Measurement method Measurement frequency | | | Acceptance inspection of general-purpose, small cells installed | Fully automated production line testing of small cells for power motors | Fully automated production line testing of large cells for xEVs or | Fully automated production line testing or large packs for xEVs or |
|--|----------------|--|---|---|--|--|
| | | | in a high-speed sorters 3561, 3561-01 | small cells for power motors or small packs of up to 60 V BT3561A | mid-size packsup of to 100 V BT3562A | large packs up of to 300 V |
| | | | 3301, 3301-01 | | | |
| | | | 33000 | NEW | NEW 33888 100 100 100 100 100 100 100 100 100 | NEW |
| | | | AC four-terminal method | AC four-terminal method | AC four-terminal method | AC four-terminal method |
| | | | 1 kHz ±0.2 Hz | 1 kHz ±0.2 Hz | 1 kHz ±0.2 Hz | 1 kHz ±0.2 Hz |
| neasurement neque | Citoy | 3 mΩ | N/A | N/A | 3.1000 mΩ, 0.1 μΩ, 100 mA | 3.1000 mΩ, 0.1 μΩ, 100 m |
| | | 30 mΩ | N/A | 31.000 mΩ, 1 μΩ, 100 mA | 31.000 mΩ, 1 μΩ, 100 mA | 31.000 mΩ, 1 μΩ, 100 mA |
| Resistance measurement ranges | | 300 mΩ | 310.00 mΩ,10 μΩ, 10 mA | 310.00 mΩ,10 μΩ, 10 mA | 310.00 mΩ,10 μΩ, 10 mA | 310.00 mΩ,10 μΩ, 10 mA |
| | | 3 Ω | 3.1000 Ω,100 μΩ, 1 mA | 3.1000 Ω,100 μΩ, 1 mA | 3.1000 Ω,100 μΩ, 1 mA | 3.1000 Ω,100 μΩ, 1 mA |
| ranges | | 30 Ω | N/A | 31.000 Ω, 1 mΩ, 100 μΑ | 31.000 Ω, 1 mΩ, 100 μΑ | 31.000 Ω, 1 mΩ, 100 μΑ |
| Max. display, | | 300 Ω | N/A | 310.00 Ω, 11 mΩ, 10 μΑ | 310.00 Ω, 10 mΩ, 10 μΑ | 310.00 Ω, 11 mΩ, 10 μΑ |
| resolution, | | 300 12 | IVA | 310.00 12, 10 π12, 10 μΑ | 310.00 Ω, 10 IIIΩ, 10 μA | 510.00 \$2, 10 111\$2, 10 μΑ |
| measurement current | | 3 kΩ | N/A | 3.1000 kΩ, 100 mΩ, 10 μA | 3.1000 kΩ, 100 mΩ, 10 μA | 3.1000 kΩ, 100 mΩ, 10 μ. |
| В | Basic | 3 mΩ range | N/A | N/A | ±0.5% rdg. ±10 dgt. | ±0.5% rdg. ±10 dgt. |
| E a | ccuracy | 30 mΩ range or more | ±0.5% rdg. ±5 dgt. | ±0.5% rdg. ±5 dgt. | ±0.5% rdg. ±5 dgt. | ±0.5% rdg. ±5 dgt. |
| | | 6 V | N/A | 6.00000 V,10 μV | 6.00000 V,10 μV | 6.00000 V, 10 μV |
| Voltage | | 20 V | 19.9999 V, 100 μV | N/A | N/A | N/A |
| measurement ranges | | 60 V | N/A | 60.0000 V, 100 μV | 60.0000 V, 100 μV | 60.0000 V, 100 μV |
| langes | | 100 V | N/A | N/A | 100.000 V, 1 mV | N/A |
| Max. display, | | 300 V | N/A | N/A | N/A | 300.000 V, 1 mV |
| resolution | | 1000 V | N/A | N/A | N/A | N/A |
| Basic accuracy | | ±0.01% rdg. ±3 dgt. *1 | ±0.01% rdg. ±3 dgt. | ±0.01% rdg. ±3 dgt. | ±0.01% rdg. ±3 dgt. | |
| Response time *2 | | | 3 ms | 10 ms | 10 ms | 10 ms |
| Sampling period '3 Ω or V Ω X:FAST, FAST, MEDIUM, SLOW Ω V | | 4 ms, 12 ms, 35 ms, 150 ms 7 ms, 23 ms, 69 ms, 252 ms | 4 ms, 12 ms, 35 ms, 150 ms 8 ms, 24 ms, 70 ms, 253 ms | 4 ms, 12 ms, 35 ms, 150 ms 8 ms, 24 ms, 70 ms, 253 ms | 4 ms, 12 ms, 35 ms, 150 m 8 ms, 24 ms, 70 ms, 253 m | |
| llowable total line re within accuracy) anges: 3 mΩ, 30 mΩ, 3 | sistance *2 *4 | SENSE line | Ν/Α, Ν/Α, 20 Ω, 20 Ω | Ν/Α, 4 Ω, 30 Ω, 30 Ω | 4 Ω, 4 Ω, 30 Ω, 30 Ω | 4 Ω, 4 Ω, 30 Ω, 30 Ω |
| | 00 mΩ, 3 Ω | SOURCE line | Ν/Α, Ν/Α, 20 Ω, 20 Ω | Ν/Α, 4 Ω, 20 Ω, 40 Ω | 4 Ω, 4 Ω, 20 Ω, 40 Ω | 4 Ω, 4 Ω, 20 Ω, 40 Ω |
| llowable total line re | sistance *2 *4 | SENSE line | Ν/Α, Ν/Α, 20 Ω, 20 Ω | Ν/Α, 6 Ω, 30 Ω, 30 Ω | 6 Ω, 6 Ω, 30 Ω, 30 Ω | 6 Ω, 6 Ω, 30 Ω, 30 Ω |
| error detection) anges: 3 mΩ, 30 mΩ, 30 | 00 mΩ, 3 Ω | SOURCE line | Ν/Α, Ν/Α, 20 Ω, 20 Ω | Ν/Α, 6 Ω, 20 Ω, 200 Ω | 6 Ω, 6 Ω, 20 Ω, 200 Ω | 6 Ω, 6 Ω, 20 Ω, 200 Ω |
| Open terminal voltage Ranges: 30 m Ω or less, 300 m Ω , 3 Ω or more | | | N/A, 7 V, 7 V peak | 25 V, 7 V, 4 V peak | 25 V, 7 V, 4 V peak | 25 V, 7 V, 4 V peak |
| LAN (TCP/IP, 10BASE-T/100BASE-TX) | | | N/A | V | ✓ | ✓ |
| RS-232C *5 (Max. 38.4 kbps) | | | ✓ (9.6 kbps fixed) | ~ | ✓ | ✓ |
| USB | | | N/A | N/A | N/A | N/A |
| GP-IB | | | ✓ (3561-01 Only) | N/A | N/A | N/A |
| EXT. I/O (37-pin Handler interface) | | | ✓ | V | ✓ | ✓ |
| Analog output (DC 0 V to 3.1 V) | | | N/A | V | ✓ | ✓ |
| Contact check | | | ✓ | ~ | ✓ | ✓ |
| Zero adjustment (±1000 counts) | | | · · | ✓ | ✓ | ✓ |
| Pulse mesurement | | | V | V | ✓ | ✓ |
| Comparator | | | Hi/ IN/ Lo | Hi/ IN/ Lo | Hi/ IN/ Lo | Hi/ IN/ Lo |
| Statistical calculations | | Max. 30,000 | Max. 30,000 | Max. 30,000 | Max. 30,000 | |
| Delay | | · · | <i>V</i> | ✓ | v | |
| Average | | | 2 to 16 times | 2 to 16 times | 2 to 16 times | 2 to 16 times |
| Panel saving/loading | | | 126 | 126 | 126 | 126 |
| Memory storage | | | 400 | 400 | 400 | 400 |
| LabVIEW® driver '6 | | | N/A Safety: EN61010 | Safety: EN61010 | Safety: EN61010 | Safety: EN61010 |
| Applicable standards Effect of radiated radio-frequency | | | EMC: EN61326 Class A | EMC: EN61326 Class A | EMC: EN61326 Class A | EMC: EN61326 Class A |
| lectromagnetic field | d (10 V/m) * | , _ | Resistant | Resistant | Resistant | Resistant |
| ffect of conducted ra | | 10 V | N/A | Resistant | Resistant | Resistant |
| requency electromagnetic field 3 V 0.15 MHz to 80 MHz, 80% AM 3 V CE | | Resistant | Resistant | Resistant | Resistant | |
| | | V | V | V | V | |
| · - | | | | | | |

DISTRIBUTED BY

Note: Company names and product names appearing in this brochure are trademarks or registered trademarks of various companies.



Test Equipment 1-800-517-8431

5 Commonwealth Ave **Depot** Woburn, MA 01801 Phone 781-665-1400 Toll Free 1-800-517-8431

Visit us at www.TestEquipmentDepot.com

^{*1:} rdg. stands for reading, dgt. stands for digits *2: Typical value *3: When the power supply frequency is 60 Hz

*4: Total line resistance = wiring resistance + contact resistance + DUT resistance *5: Available as printer I/F

*6: LabVIEW® Driver is a registered trademark of National Instruments Corporation *7: Test conditions were 80 MHz to 1 GHz at 10 V/m and 1 GHz to 6 GHz at 3 V/m, all at 80% AM

*8: Canadian Standards Assosiation