


 **Test Equipment Depot**
99 Washington St. Melrose MA, 02176
1.800.517.8431 | 781.665.0780 (Fax)
Visit us at www.TestEquipmentDepot.com

REVISIONS			
REV.	DESCRIPTIONS	DATE	APPROVED
—	CLASS A RELEASED	7/12/99	


1. ADAPTER, 6/4mm SPADE LUG WITH 4mm BANANA JACK.
2. SPECIFICATIONS:
INSULATOR: POLYACETAL
SPADE BODY: BRASS, NICKEL PLATED
3. RATING:
CURRENT: 36A MAX.
RESISTANCE: $<1\text{m}\Omega$.
TEMPERATURE: -20°C TO $+80^{\circ}\text{C}$.



PART NUMBER

CT2553-#

-COLOR: 0 — BLACK
2 — RED
4 — YELLOW
5 — GREEN
6 — BLUE
9 — WHITE

THIRD ANGLE PROJECTION		UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES TOLERANCES ARE: FRACTIONS DECIMALS ± 1/32 .XX ± .02 .XXX ±.005 DO NOT SCALE DRAWING	CONTRACT NO.		Cal Test Electronics 22820 Savi Ranch Parkway Yorba Linda, CA			
			APPROVALS	DATE	TITLE 6/4mm SPADE LUG ADAPTER w/4mm JACK			
		DRAWN	WDH	7/12/99				
		MATERIAL	CHECKED		SIZE	CAGE CODE	DWG. NO.	REV.
		FINISH	ENGR.		A	43F45	P255300	—
NEXT ASSY	USED ON		DESIGN ACTIVITY		SCALE		1:1	
APPLICATION							SHEET	1 OF 1