

Keysight E5061B Network Analyzer

NOTICE: This document contains references to Agilent Technologies. Agilent's former Test and Measurement business has become Keysight Technologies. For more information, go to **www.keysight.com**.



Notices

© Keysight Technologies 2010, 2011, 2013, 2014

No part of this manual may be reproduced in any form or by any means (including electronic storage and retrieval or translation into a foreign language) without prior agreement and written consent from Keysight Technologies as governed by United States and international copyright laws.

Manual Printing History

The manual's printing date and part number indicate its current edition. The printing date changes when a new edition is printed (minor corrections and updates that are incorporated at reprint do not cause the date to change). The manual part number changes when extensive technical changes are incorporated.

October 2010 First Edition

September 2011 Second Edition

August 2013 Third Edition

November 2014 Fourth Edition

Manual Part Number

E5061-90017

Edition

Edition 4, November 2014

Printed in Malaysia

Keysight Technologies
1-3-3 Higashikawasaki-cho, Chuo-ku Kobe
Hyogo Japan

Warranty

The material contained in this document is provided "as is," and is subject to being changed, without notice, in future editions. Further, to the maximum extent permitted by applicable law, Keysight disclaims all warranties, either express or implied, with regard to this manual and any information contained herein, including but not limited to the implied warranties of merchantability and fitness for a particular purpose. Keysight shall not be liable for errors or for incidental or consequential damages in connection with the furnishing, use, or performance of this document or of any information contained herein. Should Keysight and the user have a separate written agreement with warranty terms covering the material in this document that conflict with these terms, the warranty terms in the separate agreement shall control.

Technology Licenses

The hardware and/or software described in this document are furnished under a license and may be used or copied only in accordance with the terms of such license

Restricted Rights Legend

If software is for use in the performance of a U.S. Government prime contract or sub-contract, Software is delivered and

licensed as "Commercial computer software" as defined in DFAR 252.227-7014 (June 1995), or as a "commercial item" as defined in FAR 2.101(a) or as "Restricted computer software" as defined in FAR 52.227-19 (June 1987) or any equivalent agency regulation or contract clause. Use, duplication or disclosure of Software is subject to Keysight Technologies' standard commercial license terms, and non-DOD Departments and Agencies of the U.S. Government will receive no greater than Restricted Rights as defined in FAR 52.227-19(c)(1-2) (June 1987). U.S. Government users will receive no greater than Limited Rights as defined in FAR 52.227-14 (June 1987) or DFAR 252.227-7015 (b)(2) (November 1995), as applicable in any technical data.

Safety Notices

CAUTION

A **CAUTION** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in damage to the product or loss of important data. Do not proceed beyond a **CAUTION** notice until the indicated conditions are fully understood and met.

WARNING

A **WARNING** notice denotes a hazard. It calls attention to an operating procedure, practice, or the like that, if not correctly performed or adhered to, could result in personal injury or death. Do not proceed beyond a **WARNING** notice until the indicated conditions are fully understood and met.

Caution



Do not exceed the operating input power, voltage, and current level and signal type appropriate for the instrument being used, refer to your instrument's Function Reference.



Electrostatic discharge (ESD) can damage the highly sensitive microcircuits in your instrument. ESD damage is most likely to occur as the test fixtures are being connected or disconnected. Protect them from ESD damage by wearing a grounding strap that provides a high resistance path to ground. Alternatively, ground yourself to discharge any static charge built-up by touching the outer shell of any grounded instrument chassis before touching the test port connectors.

Safety Summary

When you notice any of the unusual conditions listed below, immediately terminate operation and disconnect the power cable.

Contact your local Keysight Technologies sales representative or authorized service company for repair of the instrument. If you continue to operate without repairing the instrument, there is a potential fire or shock hazard to the operator.

- Instrument operates abnormally.
- Instrument emits abnormal noise, smell, smoke or a spark-like light during operation.
- Instrument generates high temperature or electrical shock during operation.
- Power cable, plug, or receptacle on instrument is damaged.
- Foreign substance or liquid has fallen into the instrument.

Manufacturer's Declaration

Herstellerbescheinigung

GERA- USCHEMISSION

LpA < 70 dB

am Arbeitsplatz

normaler Betrieb

nach DIN 45635 T. 19

Manufacturer's Declaration

ACOUSTIC NOISE EMISSION

LpA < 70 dB

operator position

normal operation

per ISO 7779

Regulatory Compliance Information

This product complies with the essential requirements of the following applicable European Directives, and carries the CE marking accordingly:

- The Low Voltage Directive 2006/95/EC
- The EMC Directive 2004/108/EEC

To obtain Declaration of Conformity, please contact your local Keysight Technologies sales office, agent or distributor.

Safety Notice Supplement

- This equipment complies with EN/IEC61010-1:2001.
- This equipment is of MEASUREMENT CATEGORY I (CAT I). Do not use for CAT II, III, or IV.
- Do not connect the measuring terminals to mains.
- This equipment is a POLLUTION DEGREE 2, INDOOR USE product.
- This equipment is tested in stand-alone condition and in combination with the accessories supplied by Keysight Technologies against the requirement of the standards described in the Declaration of Conformity. If it is used as a system component, compliance of related regulations and safety requirements are to be confirmed by the builder of the system.

General Safety Precautions

The following general safety precautions must be observed during all phases of operation, service, and repair of this instrument. Failure to comply with these precautions or with specific WARNINGS elsewhere in this manual may impair the protection provided by the equipment. Such noncompliance would also violate safety standards of design, manufacture, and intended use of the instrument. Keysight Technologies assumes no liability for the customer's failure to comply with these precautions.

NOTE

The E5061B complies with INSTALLATION CATEGORY II as well as POLLUTION DEGREE 2 in IEC61010-1. The E5061B is an INDOOR USE product.

NOTE

The LEDs in the E5061B are Class 1 in accordance with IEC60825-1, CLASS 1 LED PRODUCT.

- **Ground the Instrument**

To avoid electric shock, the instrument chassis and cabinet must be grounded with the supplied power cable's grounding prong.

- **DO NOT Operate in an Explosive Atmosphere**

Do not operate the instrument in the presence of inflammable gasses or fumes. Operation of any electrical instrument in such an environment clearly constitutes a safety hazard.

- **Keep Away from Live Circuits**

Operators must not remove instrument covers. Component replacement and internal adjustments must be made by qualified maintenance personnel. Do not replace components with the power cable connected. Under certain conditions, dangerous voltage levels may remain even after the power cable has been disconnected. To avoid injuries, always disconnect the power and discharge circuits before touching them.

- **DO NOT Service or Adjust the Instrument Alone**

Do not attempt internal service or adjustment unless another person, capable of rendering first aid and resuscitation, is present.

- **DO NOT Substitute Parts or Modify the Instrument**

To avoid the danger of introducing additional hazards, do not install substitute parts or perform unauthorized modifications to the instrument. Return the instrument to a Keysight Technologies Sales and Service Office for service and repair to ensure that safety features are maintained in operational condition.

- Dangerous Procedure Warnings

Warnings, such as the example below, precede potentially dangerous procedures throughout this manual. Instructions contained in the warnings must be followed.

WARNING

Dangerous voltage levels, capable of causing death, are present in this instrument. Use extreme caution when handling, testing, and adjusting this instrument.

Safety Symbols

General definitions of safety symbols used on the instrument or in manuals are listed below.



Instruction Manual symbol: the product is marked with this symbol when it is necessary for the user to refer to the instrument manual.



Alternating current.



Direct current.



On (Supply).



Off (Supply).



In-position of push-button switch.



Out-position of push-button switch.



A chassis terminal; a connection to the instrument's chassis, which includes all exposed metal structure.



Standby.

WARNING

This warning sign denotes a hazard. It calls attention to a procedure, practice, or condition that, if not correctly performed or adhered to, could result in injury or death to personnel.

CAUTION

This Caution sign denotes a hazard. It calls attention to a procedure, practice, or condition that, if not correctly performed or adhered to, could result in damage to or destruction of part or all of the instrument.

NOTE

This Note sign denotes important information. It calls attention to a procedure, practice, or condition that is essential for the user to understand.

Certification

Keysight Technologies certifies that this product met its published specifications at the time of shipment from the factory. Keysight Technologies further certifies that its calibration measurements are traceable to the United States National Institute of Standards and Technology, to the extent allowed by the Institution's calibration facility or by the calibration facilities of other International Standards Organization members.

Documentation Warranty

The material contained in this document is provided "as is," and is subject to being changed, without notice, in future editions. Further, to the maximum extent permitted by applicable law, Keysight disclaims all warranties, either express or implied with regard to this manual and any information contained herein, including but not limited to the implied warranties of merchantability and fitness for a particular purpose. Keysight shall not be liable for errors or for incidental or consequential damages in connection with the furnishing, use, or performance of this document or any information contained herein. Should Keysight and the user have a separate written agreement with warranty terms covering the material in this document that conflict with these terms, the warranty terms in the separate agreement will control.

Exclusive Remedies

The remedies provided herein are Buyer's sole and exclusive remedies. Keysight Technologies shall not be liable for any direct, indirect, special, incidental, or consequential damages, whether based on contract, tort, or any other legal theory.

Assistance

Product maintenance agreements and other customer assistance agreements are available for Keysight Technologies products.

For any assistance, contact your nearest Keysight Technologies Sales and Service Office. Addresses are provided at the back of this manual.

Manuals for E5061B


Keysight provides the following three manuals for E5061B.

Installation Guide

The installation guide (this manual) provides start up setup information when you use the E5061B for the first time and troubleshooting information when the Windows cannot be boot up. See this manual first when you use the E5061B for the first time.

Online Help

The online help provides the information about the quick start, measurement operation, programming, built-in VBA, I/O interface This is pre-installed in the E5061B. Press **[Help]** hard key on the front panel to open. Quick Start helps in understanding the E5061B operation quickly.



The latest version of online help is available at:
<http://ena.tm.keysight.com/e5061b/manuals/webhelp/eng/>

The online help has context sensitive help, which is a great feature of the E5061B help. It allows you to get information about the selected softkey by pressing the **Help** key in the E5061B or by pressing **F1** in a keyboard attached to the E5061B or by clicking the help button in a dialog box. With context sensitive help, users can receive information quickly about the area the user is working in the firmware of the E5061B. It provides information relevant to the task that needs to be accomplished and reduces the time to search relevant information required to complete a task.

Service Guide

The service manual provides information about the parts, troubleshooting, performance test, adjustment and service menu.

In This Guide...

The following shows the contents of this manual.

Chapter 1, "Installation"

This chapter provides information about how to set up the Keysight E5061B.

Chapter 2, "Troubleshooting"

This chapter describes the troubleshooting during start up and the procedure of the operating system (OS) recovery when the Windows XP OS has been damaged.

Contents

Regulatory Compliance Information 6

Safety Notice Supplement 7

1 Installation

Checking the Shipment 18

Environmental Requirements 20

Operating Environments 20

Ventilation Requirement 20

Protection Against Electrostatic Discharge (ESD) 21

Ensuring Adequate Free Space around Analyzer for Immediate
Disconnection of Power Cable in Case of Emergency 22

Installing Front Handles/Rack Mounting Flanges 23

How to Install the Handle Kit (Option 1CN) 24

How to Install the Rack-mount Kit (Option 1CM) 25

How to Install the Rack-mount and Handle Kit (Option 1CP) 25

Connecting the Accessories 27

Connecting the BNC Adaptor (Option 1E5) 27

Connecting Mouse and Keyboard 27

Power Supply 28

Verification of the Power Supply 28

Verification and Connection of Power Cable 28

Blown Fuses 29

Starting the E5061B 30

Turning the Power ON and OFF 30

Disconnection from Supply Source 31

Initial Registration of E5061B 33

Calibration of the Touch Screen 46

2 Troubleshooting

Troubleshooting during Startup 50

System Recovery 51

Notes on executing the factory recovery function	51
Procedure to execute the factory recovery	51

Index

Regional Sales and Support Offices

1

Installation

Checking the Shipment	18
Environmental Requirements	20
Installing Front Handles/Rack Mounting Flanges	23
Connecting the Accessories	27
Power Supply	28
Blown Fuses	29
Starting the E5061B	30
Initial Registration of E5061B	33
Calibration of the Touch Screen	46

This chapter provides information about how to set up the Keysight E5061B.

Checking the Shipment

After you receive the analyzer, carry out checks during unpacking according to the following procedure.

WARNING

When unpacking the analyzer, if the exterior of the analyzer (such as the cover, front/rear panel, LCD screen, power switch, and port connectors) appear to be damaged during transport, do not turn on the power switch. Otherwise, you may get an electric shock.

Step 1. Check that the packing box or shock-absorbing material used to package the analyzer has not been damaged.

NOTE

If the packing box or shock-absorbing material has been damaged, leave the packing box and shock-absorbing material as is until other inspection items are checked as follows.

Step 2. Check the packaged items supplied with the analyzer for any damage or defects.

Step 3. By referring to the furnished contents list, check that all packaged items supplied with the analyzer have been received as per the specified options.

Step 4. After checking, if one of the following applies, contact your nearest Keysight Technologies sales and service office.

- 1 The packing box or shock-absorbing material used to package the analyzer has been damaged or the shock-absorbing material has traces where extreme force has been applied.
- 2 A packaged item supplied with the analyzer has mechanical damage or defects.
- 3 A packaged item supplied with the analyzer is missing.
- 4 A fault has been detected in the subsequent operation check of the analyzer.

If an abnormality is detected in step 1, contact the company that transported the analyzer as well as your nearest Keysight Technologies sales and service office. For

inspection by the transport company, save the packing box, shock-absorbing material, and packaged items as you received them.

Environmental Requirements

Set up the E5061B where the following environmental requirements are met.

Operating Environments

Ensure that the operating environment meets the following requirements.

Temperature	5 °C to 40 °C
Temperature range at the error-correction	23 °C ± 5 °C (< 1 °C deviation from the temperature when performing the error-correction)
Humidity	20% to 80% at wet bulb temperature < +29 °C (non-condensation)
Altitude	0 to 2,000 m (0 to 6,561 feet)
Vibration	0.21 G maximum, 5 Hz to 500 Hz

NOTE

Above environmental requirements are NOT for the specifications and measurement accuracy of the analyzer, but for the operating environment of the analyzer.

Ventilation Requirement

To ensure safety requirements, the specifications and measurement accuracy of the analyzer, you must keep the environmental temperature within the specified range by providing appropriate cooling clearance around the analyzer or, for the rack mount type, by forcefully air-cooling inside the rack housing. For more information on environmental temperature to satisfy the specifications and measurement accuracy of the analyzer, see the specification in the *E5061B Online Help*.

When the environmental temperature around the analyzer is kept within the temperature range of the operating environment specification (See Section **"Operating Environments"** on page 20), the analyzer conforms to the requirements of the safety standard. Furthermore, under that temperature requirement, the analyzer still conforms to the requirements of the safety standard even when placing the analyzer with cooling clearance as follows:

	Requirements
Back	$\leq 180 \text{ mm}$
Sides	$\leq 60 \text{ mm}$ (both right and left)




Figure 1 Ventilation space at the installation

NOTE

Place the E5061B in a proper position as show in **Figure 1**.

Protection Against Electrostatic Discharge (ESD)

Set up a static-safe work-station to protect the electronic components from the damage by the electrostatic discharge (ESD) as shown in **Figure 2**.




Figure 2 Example of the static-safe work station

Ensuring Adequate Free Space around Analyzer for Immediate Disconnection of Power Cable in Case of Emergency

As described in “**Disconnection from Supply Source**” on page 31, the power supply is disconnected by removing the power cable’s connector plug from either the AC outlet or the E5061B unit. When installing the E5061B, ensure that there is sufficient free space around the unit to permit quick disconnection of the plug (from AC outlet or E5061B unit) in case of emergency.

Installing Front Handles/Rack Mounting Flanges


The E5061B can be installed on a workbench or in a rack. This section describes how to install the front handles (Option 1CN) used for moving or transporting the instrument, and how to install the analyzer in an equipment rack as part of a measurement system (Option 1CM: without the handles, Option 1CP: with the handles).

Table 1 Keysight E5061B handles/rack mounting options

Option	Name	Keysight Part Number
1CN	Handle Kit	5063-9229
1CM	Rack-mount Kit	5063-9216
1CP	Rack-mount and Handle Kit	5188-4430

Table 2 Contents of each option

Option	Contents	Quantity
1CN	Front Handles	2
	Screws	6
	Trim Strips	2
1CM	Rack-mounting flanges (locking side plate)	2
	Screws	6
1CP	Rack-mounting flanges (locking side plate)	2
	Front Handles	2
	Screws	8



e5070aqj009

Figure 3 Installing front handle/rack-mount kits

How to Install the Handle Kit (Option 1CN)

The handle kit is used for transport and relocation of the E5061B. While referring to **Figure 3** install the handle kit by following these steps.

Step 1. Remove the adhesive-backed trim strip (1) from each side of the outer frame of the E5061B front panel.

Step 2. Use the screws provided screws to mount the front handles (2) on each side of the E5061B front panel frame.

Step 3. Attach the modified trim strip (3) provided to each front handle in order to cover the front panel locking screws.

WARNING

If the installed front handle becomes damaged, replace it with a new one immediately. A damaged handle can break while moving or lifting the instrument and cause personal injury or damage to the instrument.

How to Install the Rack-mount Kit (Option 1CM)

The rack-mount kit includes two flanges (locking side plates) for mounting the E5061B on a rack (482.6 mm/19 inches) conforming to the EIA Standard. While referring to **Figure 3**, install the rack-mount kit by following these steps.

Step 1. Remove the adhesive-backed trim strip (1) from each side of the outer frame of the E5061B front panel.

Step 2. Use the screws provided to mount a rack-mounting flange (4) on each side of the E5061B front panel frame.

Step 3. Remove the four bottom feet of the E5061B (lift the bar marked TAB on the inner side of the foot and slide the foot toward the bar).

Step 4. Mount the E5061B on the rack.

How to Install the Rack-mount and Handle Kit (Option 1CP)

The rack-mount and handle kit includes both the rack-mounting flanges (locking side plates) and front handles. While referring to **Figure 3**, install the rack-mount kit by following these steps.

Step 1. Remove the adhesive-backed trim strip (1) from each side of the outer frame of the E5061B front panel.

Step 2. Use the screws provided to mount a front handle (2) and rack-mounting flange (4) on each side of the E5061B front panel frame.

CAUTION

Use both a front handle and a rack-mounting flange in the same time certainly. Do not attempt to install flanges or handles separately with hardware provided. Serious electrical damage will occur to the instrument.

1 Installation

Step 3. Remove the four bottom feet of the E5061B (lift the bar marked TAB on the inner side of the foot and slide the foot toward the bar).

Step 4. Mount the E5061B on the rack.

Connecting the Accessories

You need to connect some accessories before use.

Connecting the BNC Adaptor (Option 1E5)

When the E5061B is equipped with Option 1E5 (high-stability frequency standard), connect the BNC adaptor from the Ref Oven terminal to the Ref In terminal on the rear panel as shown in **Figure 4**. The BNC adaptor is included in Option 1E5. When Option 1E5 is installed, the frequency accuracy and stability of the E5061B improves further.




Figure 4 Connecting the BNC adaptor

Connecting Mouse and Keyboard

The E5061B allows you to connect Mouse and/or keyboard through USB. USB mouse and keyboard can be connected with the USB ports on the front or rear panels.

Initial registration of the E5061B requires the mouse and keyboard before turning on the power.

Power Supply

Before turning on the E5061B power, check the following.

Verification of the Power Supply

Confirm that the power supplied to the E5061B meets the following requirements:

	Requirements
Voltage	90 to 132 VAC or 198 to 264 VAC ^{*1}
Frequency	47 to 63 Hz
Maximum power consumption	300 VA

* 1. Switched automatically by the E5061B in conformity to the voltage used.

Verification and Connection of Power Cable

The three-wire power cable attached to the E5061B has one wire serving as a ground. Using this power cable allows the E5061B to be grounded, thereby protecting you against electrical shock from the power outlet.

Step 1. Confirm that the power cable is not damaged.

WARNING

NEVER use a power cable showing any sign of damage. Faulty cables can cause electrical shock.

Step 2. Use the supplied cable to connect between the power cable receptacle (**Figure 5 on page 31**) on the rear panel of the E5061B and a three-wire power outlet with the grounding prong firmly connected in the ground slot.

WARNING

Use the supplied power cable with grounding wire to securely ground the E5061B.

Power cord list, 16000-99101 shows the power cable options.

Blown Fuses

If the fuse appears to have blown during operation, this instrument may be subject to failure and must be repaired. For any assistance, contact Keysight Technologies Customer contact center listed at the end of this guide.

The product uses the following fuse types:

UL/CSA Type, Slow-Blo, 10 A-250 Vac.

WARNING

DO NOT replace the fuse yourself; doing this may expose you to electrical shock.

Starting the E5061B

This section describes how to turn on/off the E5061B power and how to cut off the power supply in an emergency.

Turning the Power ON and OFF

The standby switch can turn/off the E5061B. The color on the button shows the status, as shown below:

Indicator Color	Description
Green	Normal power on status
Orange	Standby status
Red	Illegal power on status

Perform the following steps to turn the power ON or OFF.

Turning the Power ON

Step 1. Confirm if the Line Switch on the rear panel is on. The switch should always be turned on.

Step 2. Press the standby switch once and it changes to green color.

This operation turns ON the power, and the E5061B starts the self-test.

Step 3. Confirm that the self-test indicates normal operation.

Normal operation is confirmed by the self-test if no error message appears.

Turning the Power OFF

Step 1. To turn off the power of the E5061B, first, press this standby switch or send a shutdown command from the external controller to activate the shutdown process (the process of software and hardware necessary to turn off the power supply). This puts the E5061B into the standby state and the button changes to orange color.

Step 2. Next, if necessary, turn off power supply to the power cable receptacle (Line Switch) on the rear panel.

NOTE

Under normal use, **never directly interrupt the power supply to the power cable receptacle on the rear panel when the power supply is on.** Always keep the Line Switch ON and never turn it.

If you directly interrupt the power supply to the power cable receptacle when the power supply is on, or turn off the Line Switch, the shutdown process will not work. This could damage the software and hardware of the E5061B and lead to device failure.

Turning on the power supply after a faulty shutdown may cause the system to start up in a condition called "safe mode." If this occurs, first shut down the system to set it to the standby state and then turn on the power supply again to start up the system in normal mode.




Figure 5 Line switch (Always ON) and power cable receptacle

Disconnection from Supply Source

The power supply of the E5061B is cut off by disconnecting the plug of the power cable (on either AC outlet side or E5061B side). When it is necessary to disconnect the power supply in order to avoid shock hazards, etc., pull out the power cable plug from either the AC outlet side or the E5061B side.

NOTE

To allow this operation to be performed smoothly, be sure to follow the guidelines in ["Ensuring Adequate Free Space around Analyzer for Immediate Disconnection of Power Cable in Case of Emergency"](#) on page 22.

When turning the power OFF under normal circumstances, always follow the methods described in ["Turning the Power OFF"](#) on page 30.

Initial Registration of E5061B

When you start up the E5061B for the first time, you need to perform the initial registration of the Windows XP operating system of the E5061B.

NOTE

There are two kind of windows licenses for E5061B:

- Windows Vista Business
- Windows XP Pro for Embedded Systems

The license is shown at the Windows Lincese label on the E5061B rear panel.

You have different initial registration procedure depending on the windows license.

As for Windows XP Pro for Embedded Systems, you have different initial registration procedure depending on the hard disk drive (HDD revision).

NOTE

If you perform the following procedure incorrectly, a message will prompt to give you option to return to the previous registration screen and perform the registration. In such case, follow the instruction to return to the previous registration screen.

Windows XP Pro for Embedded Systems (HDD Rev KY22x and above)

Step 1. Turn on the E5061B. Do not touch any key on the front panel during boot up.

Step 2. The screen as shown in the [Figure 6](#) appears. Select **Windows XP Professional** and press **[Enter]**.




Figure 6 System start up screen in Windows XP Pro for Embedded Systems (1/2)

NOTE

After several seconds, the next screen appears automatically even if you do not press any key.

Step 3. The screen as shown in the [Figure 7](#) appears. Press **[Enter]**.




Figure 7 System start up screen in Windows XP Pro for Embedded Systems (2/2)

Step 4. The windows start up screen as shown in **Figure 8** appears.




Figure 8 Windows XP Pro for Embedded Systems start up screen

Step 5. In the Keysight Tech. Software End User License Agreement dialog box, select the **Agree** button. **Figure 9** shows the license agreement dialog box in Windows XP Pro for Embedded Systems (for HDD Rev KY22x and above).




Figure 9 Windows XP Pro for Embedded Systems license agreement dialog box

Step 6. After a while, the E5061B restarts automatically.

Windows XP Pro for Embedded Systems (HDD Rev KY21x and below)

Step 1-4. Follow the steps of Windows XP Pro for Embedded Systems (HDD Rev KY22x and above).

Step 5. In the License agreement dialog box, select the **I accept this agreement** box and click the **Next >** button. **Figure 10** shows the license agreement dialog box in Windows XP Pro for Embedded Systems (for HDD Rev KY21x and below).

1 Installation




Figure 10 Windows XP Pro for Embedded Systems license agreement dialog box

Step 6. In the next dialog box, set the date and time for your Windows. Then, click the **Next>** button (Figure 11).




Figure 11 Date and Time Setting dialog box

Step 7. After a while, the E5061B restarts automatically.

Windows Vista Business

Step 1. Follow the steps of Windows XP Pro for Embedded Systems (HDD Rev KY22x and above).

Step 2. The screen as shown in the **Figure 12** appears. Select **Microsoft Windows XP Professional** and press [Enter].




Figure 12 System start up screen in Windows Vista Business (1/2)

NOTE

After several seconds, the next screen appears automatically even if you do not press any key.

Step 3. The screen as shown in the **Figure 13** appears. Press [Enter].




Figure 13 System start up screen in Windows Vista Business (2/2)

Step 4. The Windows XP Professional Setup wizard appears. Click the **Next>** button (**Figure 14**).

1 Installation



Figure 14 Windows XP Professional Setup wizard

Step 5. In the License agreement dialog box, select the **I accept this agreement** box and click the **Next >** button.

Figure 15 shows the license agreement dialog box in Windows Vista Business.




Figure 15 Windows Vista Business license agreement dialog box

Step 6. In the next dialog box, set the date and time for your Windows. Then, click the **Next>** button (Figure 16).




Figure 16 Date and Time Setting dialog box

Step 7. After a while, the E5061B restarts automatically. The Keysight Recovery Image Setup dialog box as shown in the Figure 17 appears. Click the **OK** button.




Figure 17 Keysight Recovery Image Setup

Step 8. Once the recovery image setup is completed, the E5061B restarts again.

Calibration of the Touch Screen

When E5061B measurement screen appears, you have to calibrate the touch screen. Follow the procedure described below to calibrate the touch screen.

Step 1. Press **System** .

Step 2. Press **Service Menu**.

Step 3. Press **Test Menu**.

Step 4. Press **Adjust Touch Screen**.

The touch screen calibration screen (**Figure 18**) appears.




Figure 18 Touch Panel Calibration Screen

Step 5. Touch the symbol on the upper left of the screen with your finger. The symbols also appears on the upper right, lower left, and lower right of the screen. Touch all the symbols in that order with your finger.

Touching all the four locations described above with your finger automatically concludes the touch screen calibration.

NOTE

If you do not start to perform the touch screen calibration within several seconds, the previous measurement screen re-appears automatically.

1 Installation

2

Troubleshooting

Troubleshooting during Startup	50
System Recovery	51

This chapter describes the troubleshooting process during start up and the procedure of the operating system (OS) recovery when the Windows XP OS has been damaged.

Troubleshooting during Startup

When you encounter problems during start up, see [Table 3](#). System recovery solves most of problems.

Table 3 Troubleshooting during startup

Symptom	Solution
Turning on () the standby switch does not start up the system.	Confirm that the power cable is properly plugged in. Confirm that the line switch at the rear panel is turned on.
The system starts up, but it automatically shuts down immediately.	Execute the system recovery.
The system starts up, but it enters the service mode (The status bar at the lower right of the screen displays SVC in red).	
The measurement screen appears after start up, but the power-on test fails with Error Message 241 prompting on a red background in the instrument's message/warning area at the lower left of the screen.	

System Recovery

By executing system recovery, you can return the E5061B's system (the Windows operating system and the firmware) to the factory state (at the time of purchase^{*1}).

CAUTION

Strictly follow the steps described below. If you do any operation other than the following steps, the system may not be recovered.

Notes on executing the factory recovery function

Executing the factory recovery function causes the following conditions:

- In addition to the Windows operating system and the firmware, the following settings of the E5061B are returned to the factory state.
 - Network setting
 - GPIB setting
 - Printer setting
- The driver for the supported printer which is installed after purchase is deleted.
- You need to execute initial registration again.

Files that you have created using the save function (files in the D drive) are not affected. However, Keysight recommends backing them up before executing system recovery for precautionary purposes. For more information on backup, refer to “*Backing Up the Data*” as described in *E5061B Online Help*.

Procedure to execute the factory recovery

This section describes how to return the contents of the C drive to the factory state.

NOTE

You need a keyboard for this operation.

* 1. The default setting of the hard disk is recovered during the system recovery. For the hard disk that had failed earlier and had been replaced (after purchase), its default setting may not be the same as the factory state.

- Step 1.** Shut down the E5061B.
- Step 2.** Disconnect all of the USB device from the USB ports.
- Step 3.** Connect the keyboard to the E5061B.
- Step 4.** Press the standby switch of the E5061B to turn it on.

Windows Vista Business

- Step 5.** When the screen as shown in the figure below appears, select **Keysight Recovery System** and press [Enter].




Figure 19 Keysight Recovery System selection screen

NOTE

After several seconds, the next screen appears automatically even if you do not press any key.

Step 6. Type 2 to select **Recover Factory Backup Image** option and press [Enter].




Figure 20 System recovery selection screen

Step 7. A confirmation message appears. Type 1 and press [Enter] to continue the process.




Figure 21 System recovery confirmation screen (1/2)

Step 8. Another confirmation message appears. It also displays the approximate run time for the system configuration process. Type **1** and press **[Enter]** to continue the process. The recovery takes a few minutes depending on the amount of data.




Figure 22 System recovery confirmation screen (2/2)

Step 9. Once the recovery process is completed, completion message prompts. Press any key to continue. The E5061B restarts automatically.




Figure 23 Recovery Option confirmation dialog box

NOTE

After several seconds, the E5061B restarts automatically even if you do not press any key.

Windows XP Pro for Embedded Systems

Step 5. When the screen as shown in the figure below appears, select **Keysight Recovery System** and press [Enter].




Figure 24 Keysight Recovery System selection screen

NOTE

After several seconds, the next screen appears automatically even if you do not press any key.

Step 6. Type 2 to select **Recover Factory Backup Image** option and press [Enter].



Figure 25 System recovery selection screen

Step 7. A confirmation message appears. Click **OK** to continue the process.




Figure 26 System recovery confirmation screen

Step 8. The progress of the system configuration is displayed on the screen. The recovery takes a few minutes depending on the amount of data.




Figure 27 System configuration progress screen

Step 9. Once the recovery process is completed, completion message prompts. Click **OK**. The E5061B restarts automatically.





Figure 28 Recovery Option confirmation dialog box

Step 10. Type 5 to select **Exit and Restart the instrument** and press **[Enter]**.



Step 11. After the restart, execute initial registration. For information on the execution procedure, refer to **“Initial Registration of E5061B”** on page 33.

Step 12. Execute the calibration of the touch screen. For information on the execution procedure, refer to **“Calibration of the Touch Screen”** on page 46.

CAUTION

Never turn off the power during the system recovery because doing so may cause serious damage to the E5061B.

Index

A

accessories
 connecting accessories, 27
assistance, 11

C

connection
 connecting accessories, 27
 connecting BNC adaptor, 27
 connecting mouse and keyboard, 27

D

documentation warranty, 11

E

E5061B
 starting the E5061B, 30
environmental requirements
 operating environments, 20
 ventilation requirement, 20
error-correction
 temperature range, 20
exclusive remedies, 11

F

front handles
 how to install front handles, 23
 how to install the handle kit, 23

H

handle and rack-mount kit
 how to install the rack-mount and handle
 kit, 25
handle kit
 how to install the handle kit, 23, 24

I

initial registration, 27, 33
inspection
 checking the shipment, 18

K

Keysight Technologies sales and service office

regional sales and support offices, 61
shipping consideration due to breakdown,
 etc., 9

M

manual
 contents, 14

O

option
 Option 1CM, 23, 25
 Option 1CN, 23, 24
 Option 1CP, 23, 25
 Option 1E5, 27

P

power
 disconnection from supply source, 31
 power cable options, 28
 turning the power ON and OFF, 30
 verification and connection of power
 cable, 28
 verification of the power supply, 28
Professional Setup wizard, 41

R

rack-mount and handle kit
 how to install the rack-mount and handle
 kit, 25
rack-mount kit
 how to install the rack-mount kit, 25

S

safety
 symbols, 10
shutdown
 turning the power OFF, 30
system
 installing rack mounting flanges, 23
System Recovery
 how to execute factory recovery, 51
 how to execute user recovery function, 55,
 57
 types of system recoveries, 51

This information is subject to change without notice.
© Keysight Technologies 2010, 2011, 2013, 2014
Edition 4, November 2014

