

Test Equipment

1-800-517-8431

5 Commonwealth Ave Woburn, MA 01801 Phone 781-665-1400 Toll

GPM-8310

Digital Power Meter

FEATURES

- 5"TFT LCD
- DC, 0.1Hz ~ 100kHz Voltage/Current Test Bandwidth
- Two Numerical Display Modes
 - General Mode: Displays 2 Main Test Items + 8 Secondary Test Items
 - Simple Mode: Displays the Test Values of 4 Main Test Items
- Waveform Display: V (voltage), I (current), P (power)
- The Current/Voltage can be Measured to a Deformed Wave with CF of 3, and the Half-range CF can Reach 6 or 6A
- Meeting the IEC 61000-4-7 Harmonics Measurement Requirements (50/60Hz)
- 50th Order of Harmonic Measurement and Analysis (value and bar graph)
- Integration Function Supports Automatic Level-changing
- External Current Sensor Input Terminals (EXT1/EXT2)
- Standard Interfaces: RS-232C, USB Device/Host, LAN, GPIB
- Optional Interface: Digital I/O (DA4) (must be installed before leaving the factory)
- Optional Accessory: GPM-001



GW Instek GPM-8310 is a digital power meter for single-phase (1P/2W) AC power measurement. Features include DC, 0.1Hz~100kHz test bandwidth, 16bits A/D, and 300 kHz sampling rate. It adopts 5" TFT LCD screen with a five-digit measurement display and provides 25 power measurement related parameters, and has a high-precision measurement capability. It also features the ability to display waveform (voltage/current/power), the integration measurement function, harmonic measurement and analysis of each order (meeting the IEC 61000-4-7 harmonics measurement requirements at 50/60Hz), external sensor input terminals, and various communication interfaces, etc., to help users achieve clear, convenient and accurate power measurements. This power meter is a most cost-effective power meter with most complete functionalities among the products of the same category.

The rated direct input voltage of GPM-8310 is 600V and the input current is 20A. The minimum current level is 5mA (resolution up to 0.1uA) and the power measurement resolution is 0.1uW. The crest factor can reach 3 (half measurement range can reach 6 or 6A), and the voltage/current/power measurement capability can reach (±0.05% reading ±0.1% level). Different measurement modes can be selected according to (AC+DC/AC/DC/V-MEAN), providing up to 25 relevant parameters for power measurement, including voltage (Vrms/Vac/Vdc/Vmn/V+pk/V-pk), current (Irms/Iac/Idc/I+pk/I-pk), frequency (VHz/IHz), power (P/P+pk/P-pk), crest factor (CFV/CFI), apparent power (VA), reactive power (VAR), power factor (PF), phase angle (DEG), total harmonic distortion rate (THDV/THDI), maximum current ratio (MCR), and the MATH calculation function. Hence, for the measurement of low current/low power such as standby power consumption, or the measurement of power consumption of general products, this power meter provides the best range and accuracy support.

GPM-8310 also makes good use of the advantages of the TFT LCD to display the results of parameter measurement by using numerical and graphical methods. In terms of numerical values, the general mode and the simple mode are provided. The general mode can display 10 measurement parameters (2 main measurements + 8 monitoring measurements), and the simple mode can display four measurement parameters. These displayed parameters can be arbitrarily selected from 25 power parameters according to the needs of users. In terms of graphic display, a simple oscilloscope mode is provided to display waveforms for three parameters including voltage, current and power. In addition, the measurement and analysis of each harmonic order of the measurement signal can be completely displayed by numerical values or bar graphs. This power meter not only meets the needs of accuracy and legibility in process testing, but also meets the needs of diverse measurement applications in R&D design and quality verification.

In addition, the performance of GPM-8310 in auxiliary measurement mechanism/function is also comprehensive. For the application of measuring large voltage, the VT rate setting can be used with an external voltage Potential Transformer. For the measurement of large current, the type of current transformer ~ voltage output type or current output type will determine the applied method. If it is a current output type, it can be directly locked to the rear panel of the instrument and collocated with the CT rate setting to conduct measurement. If it is a voltage output type, measurement can be conducted through the external current sensor input terminals (EXT1/EXT2) provided by GPM-8310. Automatic level-changing can self-define the required level to save level-changing time. 10,000 lots of internal memories can be used to store measurement data according to the update rate set by GPM-8310 or a user-defined time interval for subsequent analysis.

In terms of data retrieval and storage, GPM-8310 provides a variety of communication interfaces including RS-232C/ USB device (virtual COM)/ LAN/ GPIB. Users can write programs to read the measurement results according to their habits or with existing system interfaces and there is no need to procure interfaces. USB host supports GPM-8310 screen capture, internal record data access, and firmware update. For the needs of external signal control or the use of data recorder to record data, GPM-8310 also provides an optional Digital I/O (DA4) interface (must be installed before leaving the factory), which can be connected to an external controller such as PLC or a data recorder to meet the application of automatic measurement or long recording.

VARIOUS DISPLAY MODES











Numerical (General) Mode Numerical (Simple)
Mode

Waveform Mode Harmonic (Bar Graph)
Measurement

Harmonic (Table Column)

Measurement

GPM-8310 provides the numerical value display mode and the waveform display mode, which help users to maximize the benefit of their measurement. Under the numerical mode, there are the general mode and the simple mode. The general mode has related measurement settings and can simultaneously display 10 measurement parameters (2 main measurements and 8 secondary measurements). The simple mode displays only 4 measurement parameter results. The parameters in each mode can be arranged and combined as required. Under the graphic mode, a simple oscilloscope function is provided to display the waveforms of three parameters including voltage, current and power. The horizontal

scale can be adjusted (from 25us/div ~ 1s/div according to the set data update rate), and 3 magnification rates for waveform observation are also provided for users to select. In the harmonic measurement, the measurement results of each order of harmonics can be displayed by bar graphs, and a specific observation order can be specified. The relevant values of each order of harmonics (voltage/current/power/voltage distortion ratio/current distortion ratio/power distortion ratio/voltage phase angle/current phase angle) can be completely recorded and displayed.

Measurement Items	Symbols
Voltage	Vrms, V+pk, V-pk, Vac*, Vdc*, Vmn*
Current	Irms, I+pk, I-pk, Iac*, Idc*
Power	P, P+pk, P-pk, VA, VAR
Power Factor	PF
Crest Factor	CFV, CFI
Phase Angle	DEG
Frequency	VHz, IHz
Total Harmonic Distortion	THDV, THDI
Maximum Current Ratio	MCR
Integration	WP, WP+, WP-, q, q+, q-, Vac, Iac

Note: "*" Only applicable to specific measurement modes for selection

GPM-8310 provides a variety of measurement items and functions, including voltage, current, frequency, effective power, apparent power, reactive power, power factor, crest factor, total harmonic distortion, and can also measure the maximum current ratio. GPM-8310 is also equipped with the measurement function of power or current time integration for the DUT. Users set a period of time to perform instantaneous power


integration at the set time period, and then divide by the time to obtain the average power of the DUT. In addition, when performing integration measurement, GPM-8310 supports automatic level-changing function for the power change of the DUT at different times in order to obtain the most complete integration result within the set time.

C. SUPERB MEASUREMENT ASSISTANCE



Ratio Configuration

With respect to the support of measurement assistance, the performance of GPM-8310 is outstanding. First of all, for the measurement of high voltage/high power, the setting of voltage ratio/power ratio is provided to restore the attenuated ratio to a true value. For the measurement of large current, other than the setting of current ratio, external current sensor terminals (EXT1/EXT2) can be utilized to connect with a voltage output type current transformer, making large current measurement more



External Current Sensor Input

convenient. In addition, GPM-8310 provides 4 sets of panel settings for storage/recall and memory for storing 10,000 lots of measurement values. The measurement storage can log the measurement results based upon the update rate or a self-defined time interval to facilitate the subsequent analysis. The USB host on the front panel supports screen capture, measurement value storage, and GPM-8310 firmware update.

D. FLEXIBLE LEVEL-CHANGING MECHANISM



Automatic level-changing under the integration function

GPM-8310 provides the measurement of the integration function under the automatic level-changing mode to allow users to fully calculate the total value of the power consumption of the DUT from the beginning to the end of the integration function. In addition, GPM-8310 also supports




Self-defined automatic level-changing mechanism

self-defined setting mechanism for level-changing. Users can select the required level to be changed to save time on level-changing and expedite the test.



Practical Interface

GPM-8310 provides comprehensive and diverse communications interfaces including RS-232 / USB / LAN / GPIB, which are suitable for customers to write computer software for remote control and the collection of measurement results through commands. The optional Digital I/O (DA4) interface provides 3 different modes: the external control mode, the DA4 output mode and the self-defined output mode based on user settings. When the setting is in the external control mode, it allows users to activate, stop, trigger or reset the integration measurement



DA4 Interface Mechanism

function through external signals. When the setting is in the DA4 output mode, users can define 4 measurement parameter values from the 25 measurement parameters provided (even with the result of integration measurement) to produce outputs by a fixed level (full scale $\pm5V$) or a manual level (full scale $\pm5V$) and receive results by collocating with a data recorder. When the setting is in the self-defined output mode, a communications interface is required to control the action of each defined pin through commands.

PANEL INTRODUCTION









GTL-210 Test Lead



GPM-001 Test Fixture



GPM-001(EU) Test Fixture

Input Type	SPECIFICATIONS					
Measure Range	INPUT					
Current Flouring, injust through shust	Item					
Curreint Small Toma, 20ma, 50ma, 100ma, 20ma, 0.5a, 1.a, 2a, 5a, 10.a, 20.a Sensor input ERT 1.2.5 x y 1.0 v 20 mt, 500 mt, 20 mt, 500 mt, 1.9, 2 y 100 pt resistance approach 2 Mts 100 pt resistance approach 50 fms 20 pt resistance 20 pt resistanc	Input Type					
Input Impedance	Measure Range	Current Direct input 5mA, 10mA, 20mA, 50mA, 100mA, 200mA, 0.5A, 1A, 2A, 5A, 10A, 20A Sensor input EXT 1: 2.5 V, 5 V, 10 V				
Current Direct input range SNA - 200m Direct imput rang						
Direct input range 5mA – 200m A prior trigingt are go. 3A – 20A Sensor input input range 25 v – 10V (EXT1) Input range 25 v – 10V (EXT2) Input range 25 v –	Input Impedance			Input resistance: ap	proach 2 M Ω	
Continuous Maximum Allowable Input Voltage Current Direct input range 5m > 200mA Direct range 1m > 200mA D		Direct input ra Direct input ra Sensor input Input range 2.5	nge 0.5A ~ 20A 5V ~ 10V (EXT1)	Input resistance: ap	proach 5 m Ω	
Currient Direct input range 0 5M = 200mA Sensor input Sensor			mV ~ 2V (EXT2)		•	
Continuous Maximum Common-mode Voltage Line Filter select OFF or ON (cut off frequency of 500 Hz) select OFF or ON (cut off frequency of 500 Hz) Simultaneous conversion voltage and current inputs Resolution 16bits Maximum conversion rate Approx. 300kHz VOLTAGE AND CURRENT ACCURACY Item Requirements Temperature 23 ± 5°C Humidility 30.75% RH Input waveform Sine wave crest factor = 3 common-mode voltage 0 V Number of displayed digits 5 digits Frequency filter Turn on to measure voltage or current of 200 Hz or less Afer 30 minutes after warm-up time has passed Afer measurement range is changed (zero-level compensation) Update interval is 250 ms DC 1Hz ≤ f < 45 Hz 10.1% of reading + 0.2% of range) 45 Hz ≤ f < 66 Hz 10.1% of reading + 0.2% of range) 45 Hz < f < 100 kHz < f < 100 kHz < f < 100 kHz 1 kHz < f < 100 kHz < f < 100 kHz 45 + 66 Hz 40.2% of reading + 0.2% of range) 10 kHz < f < 100 kHz < f < 100 kHz 45 + 66 Hz 45 + 66 Hz 46 0.2% of reading + 0.5% of range) 1 (0.04x(f-10))% of reading 100 kHz 45 + 66 Hz 46 0.2% of reading + 0.5% of range) 1 (0.04x(f-10))% of reading 100 kHz 45 + 60 Hz Accuracy When the Creet Factor is Set to 6 or 6A Accuracy Vhen the Creet Factor is Set to 6 or 6A Accuracy Changes Caused by Data Update interval is 100 ms, and Auto, add 0.05% of reading to the 0.1 Hz to 1 kHz accuracy. Add the following value to the DC current accuraces. S mA/10 mA/20 mA/20 mA/20 mA/20 mA ranges S mA/10 mA/20 mA/20 mA/20 mA ranges S mA/10 mA/20 mA/20 mA/20 mA ranges External current sensor input (EXT1) 1 miV/C 2 maccuracy When the Creet Factor is Set to 6 or 6A Accuracy When the Creet Factor is Set to 6 or 6A Accuracy When the Creet Factor is Set to 6 or 6A Accuracy Changes Caused by Data Update Interval is 100 ms, and Auto, add 0.05% of reading to the 0.1 Hz to 1 kHz accuracy. Add the following value to the DC current accuraces. S mA/10 mA/20 mA/2	Continuous Maximum Allowable Input	Current Direct input ra Direct input ra		A peak value of 30 A o peak value of 100A o	or RMS value of 20A, whichever is less or RMS value of 30A, whichever is less	
Select OFF or ON (cut off frequency of 500 Hz)	Input Bandwidth	DC, 0.1 Hz ~ 1	00kHz			
Select OFF or ON (cut off frequency of 500 Hz)		600 Vrms, CAT	П			
Simultaneous conversion voltage and current inputs	Line Filter	select OFF or 0	ON (cut off frequ	ency of 500 Hz)		
Resolution 16bits Maximum conversion rate Approx. 300kHz	• •					
Specifications	A/D Converter	Resolution 16b	oits			
Requirements Temperature 23 ± 5 °C Humidity 30.75 % RH Input waveform Sine wave crest factor = 3 0.75 % RH Input waveform OV Number of displayed digits 5 digits Frequency filter After 30 minutes after warm-up time has passed After measurement range is changed (zero-level compensation) Update interval is 250 ms 25	VOLTAGE AND CURRENT ACCURACY					
Humidity Input waveform Sine wave crest factor = 3 common-mode voltage Number of displayed digits Frequency filter After 30 minutes after warm-up time has passed After measurement range is changed (zero-level compensation) Update interval is 250 ms Accuracy C C ± (0.1% of reading + 0.2% of range) 10.1 Hz ≤ f < 45 Hz ± (0.1% of reading + 0.0% of range) 10.1 Hz ≤ f < 60 Hz 45 Hz ≤ f ≤ 60 Hz 10.1 Hz ≤ f < 10 Hz 10.1		Specifications				
$0.1 \text{Hz} \le f < 45 \text{Hz} \qquad \pm (0.1 \% \text{of reading} + 0.2 \% \text{of range})$ $45 \text{Hz} \le f \le 66 \text{Hz} \qquad \pm (0.1 \% \text{of reading} + 0.05 \% \text{of range})$ $16 \text{Hz} \le f \le (0.1 \% \text{of reading} + 0.3 \% \text{of range})$ $16 \text{Hz} \le f \le (0.1 \% \text{of reading} + 0.3 \% \text{of range})$ $16 \text{Hz} \le f \le 10 \text{kHz} \qquad \pm (0.07 \text{\%f}) \% \text{of reading} + 0.3 \% \text{of range})$ $10 \text{kHz} < f \le 100 \text{kHz} \qquad \pm (0.5 \% \text{of reading} + 0.3 \% \text{of range}) \pm [(0.04x(f-10))\% \text{of reading}]$ $Add \qquad \pm 0.03\% \text{of reading} + 0.5 \% \text{of range}) \pm [(0.04x(f-10))\% \text{of reading}]$ $Add \qquad \pm 0.03\% \text{of reading} + 0.2 \% \text{of range}) \pm [(0.04x(f-10))\% \text{of reading}]$ $Add \qquad \pm 0.03\% \text{of reading} + 0.2 \% \text{of range}) \pm (0.04x(f-10))\% \text{of reading}]$ $Accuracy \text{When the Line Filter is Turned ON}$ $Accuracy \text{When the Crest Factor is Set to 6 or 6A}$ $Accuracy \text{Changes Caused by Data}$ $Update Interval \text{Influence of Temperature Changes}$ $After \text{Zero-level Compensation or}$ $Range \text{Change}$ $Add 0.02\% \text{of range}/^{x}C \text{to the DC voltage accuracy.}$ $Add 0.02\% \text{of range}/^{x}C \text{to the DC current accuracies.}$ $5 \text{mA}/10 \text{mA}/20 \text{mA}/300 \text{mA}/200 \text{mA ranges}$ $5 \text{mA}/^{x}C$ $0.5 \text{A}/ \text{A}/ $	Requirements	Humidity Input waveforn common-mode Number of dis Frequency filte After 30 minute After measurer	e voltage played digits r es after warm-up nent range is cha	30~75% RH Sine wave crest factor 0 V 5 digits Turn on to measure vo	oltage or current of 200 Hz or less	
Temperature Coefficient Add ±0.03% of reading/"C within the range 5 to 18°C or 28 to 40°C. When the Line Flter is Turned ON 45 − 66 Hz Add 0.2 % of reading Accuracy When the Crest Factor is Sset to 6 or 6A Accuracy Changes Caused by Data Update Interval Influence of Power Factor Add 0.02% of range/"C to the DC voltage accuracy. Add 0.02% of range/"C to the DC voltage accuracy. Add the following value to the DC current accuracies. 5 mA/10 mA/20 mA/50 mA/100 mA/200 mA ranges 5 pA/"C 0.5 A/1 A/2 A/5 A/10 A/20 A ranges 10 mV/"C External current sensor input (/EXT1) External current sensor input (/EXT2) The standard outside the crest factor is set to 6 or 6A Accuracy Changes Caused by Data Update Interval ACTIVE POWER ACCURACY Item Specifications Requirements DC 0.1 Hz ≤f≤45 Hz ± (0.3 % of reading + 0.2 % of range) 45 Hz ≤f≤66 Hz ± (0.1 % of reading + 0.2 % of range) 10 kHz <f≤ %="" (s:="" (λ)="0" +="" 0.5="" 10="" 100="" 40.0="" <f≤="" [{0.067x(f-1)}%="" [{0.09x(f-10)}%="" apparent="" by="" crest="" factor="" influence="" is="" khz="" of="" power="" power)="" power)<="" range="" range)="" reading="" reading}="" set="" td="" the="" to="" when="" ±=""><td>Accuracy</td><td>0.1 Hz \leq f$<$ 45 45 Hz \leq f\leq 66 66 Hz$<$ f\leq 1 k 1 kHz$<$ f\leq 10</td><td>Hz $\pm (0)$ Hz $\pm (0)$ Hz $\pm (0)$ Hz $\pm (0)$ $\pm (0)$</td><td>.1 % of reading + 0.2 % o .1 % of reading + 0.05 % o .1 % of reading + 0.2 % o .07 *f) % of reading + 0.3</td><td>f range) of range) f range) 1% of range)</td></f≤>	Accuracy	0.1 Hz \leq f $<$ 45 45 Hz \leq f \leq 66 66 Hz $<$ f \leq 1 k 1 kHz $<$ f \leq 10	Hz $\pm (0)$ Hz $\pm (0)$ Hz $\pm (0)$ Hz $\pm (0)$ $\pm (0)$.1 % of reading + 0.2 % o .1 % of reading + 0.05 % o .1 % of reading + 0.2 % o .07 *f) % of reading + 0.3	f range) of range) f range) 1% of range)	
When the Line Fiter is Turned ON 45 ~ 66 Hz	Temperature Coefficient					
Accuracy When the Crest Factor is Set to 6 or 6A Accuracy Changes Caused by Data Update Interval Influence of Temperature Changes After Zero-level Compensation or Range Change Accuracy When the Crest Factor is Set to 6 or 6A Accuracy Changes Caused by Data Update Interval Influence of Temperature Changes After Zero-level Compensation or Range Change Add 0.02% of range/°C to the DC voltage accuracy. Add the following value to the DC current accuracies. 5 mA/10 mA/20 mA/50 mA/100 mA/200 mA ranges 5 mA/°C 0.5 A/1 A/2 A/5 A/10 A/20 A/20 A ranges 5 mA/°C 0.5 A/1 A/2 A/5 A/10 A/20 A ranges 5 mA/°C 0.5 A/1 A/2 A/5 A/10 A/20 A ranges 5 mA/°C 1 mV/°C 2 accuracy When the Crest Factor is Set to 6 or 6A Accuracy Changes Caused by Data When the data update interval is 100 ms, and Auto, add 0.05% of reading to the 0.1 Hz to 1 kHz accuracy. When the data update interval is 100 ms, and Auto, add 0.05% of reading to the 0.1 Hz to 1 kHz accuracy. When the data update interval is 100 ms, and Auto, add 0.05% of reading to the 0.1 Hz to 1 kHz accuracy. When the data update interval is 100 ms, and Auto, add 0.05% of reading to the 0.1 Hz to 1 kHz accuracy. Accuracy Changes Caused by Data When the data update interval is 100 ms, and Auto, add 0.05% of reading to the 0.1 Hz to 1 kHz accuracy. When the data update interval is 100 ms, and Auto, add 0.05% of reading to the 0.1 Hz to 1 kHz accuracy. Accuracy Item Specifications same as the conditions for voltage and current. Power factor 1 DC (0.1 % of reading + 0.2 % of range) 0.1 Hz \leq f \leq 45 Hz 1 kHz \leq f \leq 66 Hz 1 kHz \leq f \leq 66 Hz 1 kHz \leq f \leq 10 kHz 1 kHz \leq for feading + 0.2 % of range) 1 kHz \leq f \leq 10 kHz 1 kHz \leq for feading + 0.5 % of range) 1 kHz \leq f \leq 10 kHz 2 k(0.1 % of reading + 0.5 % of range) 1 kHz \leq f \leq 10 kHz 2 k(0.1 % of reading + 0.5 % of range) 1 kHz \leq f \leq 10 kHz 2 k(0.1 % of reading + 0.5 % of range) 1 kHz \leq f \leq 10 kHz 2 k(0.1 % of	-		Add	0.2 % of reading	C	
Update Interval Influence of Temperature Changes After Zero-level Compensation or Range Change Add 0.02% of range/°C to the DC voltage accuracy. Add the following value to the DC current accuracies. 5 mA/10 mA/20 mA/20 mA/100 mA/200 mA ranges 5 $0.0 \mu A/^{\circ}C$ 0.5 A/1 A/2 A/5 A/10 A/20 A ranges 500 $\mu A/^{\circ}C$ 20 External current sensor input (/EXT1) 1 mV/°C 2 2 2 2 3 $0.0 \mu A/^{\circ}C$ 3 accuracy When the Crest Factor is Set to 6 or 6A 2 2 2 2 3 $0.0 \mu A/^{\circ}C$ 3 accuracy obtained by doubling the measurement range error for the accuracy when the crest factor is set to 6 or 6A 3 2 2 2 2 2 2 2 2 2 2 2 2 2 2 2 2 2 2	Accuracy When the Crest Factor is Sset to 6 or 6A	accuracy obtain		· ·	rror for the accuracy when the crest factor is set to	
Influence of Temperature Changes After Zero-level Compensation or Range Change Add 0.02% of range/°C to the DC voltage accuracy. Add the following value to the DC current accuracies. 5 mA/10 mA/20 mA/20 mA/20 mA ranges 5 mA/°C 0.5 A/1 A/2 A/5 A/10 A/20 A ranges 5 00 μA/°C External current sensor input (/EXT1) External current sensor input (/EXT2) 50 μV/°C Accuracy When the Crest Factor is Set to 6 or 6A Accuracy Changes Caused by Data Update Interval ACTIVE POWER ACCURACY Item Specifications same as the conditions for voltage and current. Power factor 1 Accuracy DC 0.1 % of reading + 0.2 % of range) 4.5 Hz ≤ f < 45 Hz 4 (0.3 % of reading + 0.2 % of range) 6.6 Hz < f ≤ 1 kHz 1 kHz < f ≤ 10 kHz 1 kHz vertactor When power factor (λ) = 0 (S: apparent power)	Accuracy Changes Caused by Data Update Interval	When the data	update interval is	s 100 ms, and Auto, add (0.05% of reading to the 0.1 Hz to 1 kHz accuracy.	
to 6 or 6AAccuracy Changes Caused by DataWhen the data update interval is 100 ms, and Auto, add 0.05% of reading to the 0.1 Hz to 1 kHz accuracy.Update IntervalSpecificationsRequirementsSpecifications same as the conditions for voltage and current. Power factor 1AccuracyDC $(0.1\% \text{ of reading} + 0.2\% \text{ of range})$ $45 \text{ Hz} \le f < 45 \text{ Hz}$ $\pm (0.3\% \text{ of reading} + 0.2\% \text{ of range})$ $45 \text{ Hz} \le f \le 66 \text{ Hz}$ $\pm (0.1\% \text{ of reading} + 0.2\% \text{ of range})$ $45 \text{ Hz} \le f \le 16 \text{ Hz}$ $\pm (0.2\% \text{ of reading} + 0.2\% \text{ of range})$ $166 \text{ Hz} < f \le 10 \text{ kHz}$ $\pm (0.2\% \text{ of reading} + 0.3\% \text{ of range}) \pm [\{0.067x(f-1)\}\% \text{ of reading}]$ Influence of Power Factorwhen power factor $(\lambda) = 0$ (S: apparent power)	Influence of Temperature Changes After Zero-level Compensation or Range Change	Add the followi 5 mA/10 mA/2 0.5 A/1 A/2 A/ External curren	ing value to the D 0 mA/50 mA/100 5 A/10 A/20 A rai it sensor input (/	OC current accuracies. O mA/200 mA ranges nges EXT1)	500 µA/°C 1 mV/°C	
$ \begin{tabular}{l lllllllllllllllllllllllllllllllllll$	Accuracy When the Crest Factor is Set to 6 or 6A	accuracy obtain	ned by doubling t	he measurement range e	rror for the accuracy when the crest factor is set to	
Specifications Specifications Same as the conditions for voltage and current. Power factor 1	Accuracy Changes Caused by Data Update Interval	When the data	update interval i	s 100 ms, and Auto, add (0.05% of reading to the 0.1 Hz to 1 kHz accuracy.	
Requirements same as the conditions for voltage and current. Power factor 1 Accuracy DC $(0.1 \% \text{ of reading} + 0.2 \% \text{ of range})$ $45 \text{ Hz} \le f < 45 \text{ Hz}$ $\pm (0.3 \% \text{ of reading} + 0.2 \% \text{ of range})$ $45 \text{ Hz} \le f \le 66 \text{ Hz}$ $\pm (0.1 \% \text{ of reading} + 0.05 \% \text{ of range})$ $66 \text{ Hz} < f \le 10 \text{ kHz}$ $\pm (0.2 \% \text{ of reading} + 0.2 \% \text{ of range})$ $1 \text{ kHz} < f \le 10 \text{ kHz}$ $\pm (0.1 \% \text{ of reading} + 0.3 \% \text{ of range}) \pm [\{0.067x(f-1)\}\% \text{ of reading}]$ Influence of Power Factor when power factor (λ) = 0 (S: apparent power)	ACTIVE POWER ACCURACY	Specifications				
Accuracy $ \begin{array}{ccccccccccccccccccccccccccccccccccc$	Requirements	same as the co	onditions for volta	age and current.		
	Accuracy	DC $0.1Hz \le f < 45$ $45Hz \le f \le 66$ $66Hz < f \le 1k$ $1kHz < f \le 10$	Hz $\pm (0)$.3 % of reading + 0.2 % o .1 % of reading + 0.05 % .2 % of reading + 0.2 % o .1 % of reading + 0.3 % o	of range) of range) of range) of range) of range) of range) ± [{0.067x(f-1)}% of reading]	
$\pm 0.1 \%$ of S for 45 Hz \leq f \leq 66 Hz	Influence of Power Factor					

SPECIFICATIONS						
	$\pm \{(0.1 + 0.15 \times f) \% \text{ of S} \}$ for up to 100 kHz as reference data					
	•f is frequency of input signal in kHz when $0 < \lambda < 1$ (Φ : phase angle of the Voltage and current)					
	(power reading) \times [(power reading error%) + (power range %) \times (power range / indicated apparent					
NA -1 1 -1 -1 -1 -1 -1 -1 -1 -1 -1 -1 -1 -	power value) + $\{\tan \Phi \times (\text{influence when } \lambda = 0)\%\}$					
When The Line Filter is Turned ON	45 ~ 66 Hz Add 0.3 % of reading < 45 Hz Add 1 % of reading					
Temperature Coefficient	same as the temperature coefficient for voltage and current					
Accuracy When The Crest Factor is Set to 6 or 6A	accuracy obtained by doubling the measurement range error for the accuracy when the crest factor is set to 3					
Accuracy of Apparent Power S	voltage accuracy + current accuracy					
Accuracy of Reactive Power Q	accuracy of apparent power + $(\sqrt{1.0004 - \lambda 2}) - (\sqrt{1 - \lambda 2}) \times 100 \%$					
Accuracy of Power Factor Λ	$\pm [(\lambda - \lambda/1.0002) + [\lambda \cos \theta - \cos \{\theta + \sin \theta] $ (influence from the power factor when $\lambda = 0\%/100)\}$ [] ± 1 digit					
Accuracy of Phase Difference A	when voltage and current are at the measurement range rated input $\pm [[\phi - \cos^{-1}(\lambda/1.0002)] + \sin^{-1}(influence from the power factor when \lambda = 0 \% / 100)] \pm 1 digit$					
Accuracy of Phase Difference Φ	when voltage and current are at the measurement range rated input					
Accuracy When The Crest Factor is Set to 6 or 6A	accuracy obtained by doubling the measurement range error for the accuracy when the crest factor is set to 3					
Accuracy Changes Caused by Data Update Interval	When the data update interval is 100 ms, and Auto, add 0.05% of reading to the 0.1 Hz to 1 kHz accuracy.					
VOLTAGE, CURRENT AND ACTIVE PO	· · · · · · · · · · · · · · · · · · ·					
Item	Specifications					
Measurement Method	Digital sampling method					
Crest Factor Wiring System	3 or 6 (6A) Single-phase, two-wire (1 P2 W)					
Range Select	Select manual or auto ranging					
Auto Range	Auto-range increase					
	The range is upped when any of the following conditions is met. Crest factor 3 Urms or Irms exceeds 130% of the currently set measurement range.					
	Upk, Ipk value of the input signal exceeds 300% of the currently set measurement					
	range. Crest factor 6 Urms or Irms exceeds 130% of the currently set measurement range.					
	Upk, Ipk value of the input signal exceeds 600% of the currently set measurement					
	range. Crest factor 6A Urms or Irms exceeds 260% of the currently set measurement range.					
	Upk, Ipk value of the input signal exceeds 600% of the currently set measurement range.					
	Auto-range decline					
	The range is downed when all of the following conditions are met.					
	Crest factor 3 Urms or Irms is less than or equal to 30% of the measurement range. Urms or Irms is less than or equal to 125% of the next lower measurement range.					
	Upk, Ipk value of the input signal exceeds 300% of the currently set					
	measurement range. Crest factor 6 or 6A Urms or Irms is less than or equal to 30% of the measurement range.					
	Urms or Irms is less than or equal to 125% of the next lower measurement range.					
	Upk, Ipk value of the input signal exceeds 600% of the currently set measurement range.					
Display Mode Switching	Vrms (the true RMS value of voltage and current)					
	VOLTAGE MEAN (the rectified mean value calibrated to the RMS value of the voltage and the true RMS value of the current)					
	AC ,					
Marriago de 1 de 10	DC					
Measurement Synchronization Source	Select voltage, current, or off In the case of Auto Update Rate, select the voltage or current from the equipped element.					
Line Filter	Select OFF or ON (cutoff frequency at 500 Hz).					
Peak Measurement	Measures the peak (max, min) value of voltage, current or power from the instantaneous voltage, instantaneous current or instantaneous power that is sampled.					
Zero-level Compensation	Removes the internal offset of the measure unit (After measurement range is changed)					
Measurement Parametersl	Voltage Vrms , Vmn, Vdc , Vac					
	Current Irms , Idc , Iac Active Power P					
	Apparent Power VA					
	Reactive power VAR Power Factor PF					
	Crest Factor CFI, CFV					
	Phase Angle DEG					
	Frequency IHz and VHz Voltage Peak V+pk and V-pk					
	Current Peak I+pk and I-pk					
	Active Power Peak P+pk and P-pk					
	Total Harmonic Distortion THDI and THDV Maximum Current Ratio MCR					
	1					

SPECIFICATIONS						
FREQUENCY MEASUREMENT						
Item	Specifications					
Measurement Item	Voltage and current					
Measurement Frequency Range	Data update interval 0.1 s 0.25 s 0.5 s 1 s 2 s 5 s 10 s 20 s Auto (*) (*) Limit of the measure Timeout 1 s 5 s 10 s 20 s	Measurement Freque 20 Hz \leq f \leq 100 kHz 10 Hz \leq f \leq 100 kHz 5 Hz \leq f \leq 100 kHz 2.0 Hz \leq f \leq 100 kHz 1.0 Hz \leq f \leq 100 kHz 1.0 Hz \leq f \leq 100 kHz 0.5 Hz \leq f \leq 100 kHz 0.2 Hz \leq f \leq 100 kHz 0.1 Hz \leq 0.1 Hz 0.2 Hz 0.3 Hz	uency by the Timeou	t setting		
Measurement Range	Auto switching among si	x types: 100mHz, 1 H	z, 10 Hz, 100 Hz, 1 k	Hz, 10 kHz, and 100 kHz.		
Frequency Filter	Select OFF or ON (cut of	ff frequency of 500 Hz	:)			
Accuracy	Requirements When the input signal level is 30% or more of the measurement range If the crest factor is set to 3. (60% or more if the crest factor is set to 6 or 6A) • Frequency filter is ON when measuring voltage or current of 200 Hz or less.					
INTEGRATION	± (0.06% of reading)					
Item	Specifications					
Mode	Select manual integration	n mode, standard inte	gration mode, or rep	etitive integration mode.		
Timer	Automatically stop integration by setting a timer. Selectable range: 0 hours 00 minutes 00 seconds to 9999 hours 59 minutes 59 seconds					
Accuracy	±(Power accuracy (or cur			nge)		
Range Setting	Auto range or fixed range ±0.02%	e is available for integ	ration			
Timer Accuracy Remote Control	Start, stop and reset ope	rations are available u	ısing an external rem	ote signal (option)		
HARMONIC MEASUREMENT	otart, stop and reset spe		ionig un onternur reni	ote orginali (option)		
Item	Specifications					
Measured Item	Voltage, Current, Power					
Measured Method	Zero-cross simultaneous	calculation method				
Frequency Range FFT Data Length	10 Hz to 1.2 kHz. 1024					
•	4096 (Auto switch when	both 50Hz/60Hz and	update rate > 0.1s co			
Sample Rate, Window Width, and Upper Limit of Analysis Orders*	Fundamental Frequency 10 Hz to 44 Hz 45 Hz to 55 Hz 54 Hz to 66Hz 67 Hz to 150 Hz 150 Hz to 300 Hz 300 Hz to 600 Hz 600 Hz to 1200 Hz	Sample rate f × 1024 f x 512 f x 512 f × 512 f × 256 f × 128 f × 64	Window Width 1 10 12 2 4 8 16	upper limit of Analysis orders 50 50 50 32 16 8		
Accuracy	Frequency $10 \text{ Hz} \le f < 45 \text{ Hz}$ $45 \text{ Hz} \le f < 440 \text{ Hz}$ $440 \text{ Hz} \le f < 1.2 \text{kHz}$	Voltage 0.15% of reading + 0.35% of range 0.15% of reading + 0.35% of range 0.20% of reading + 0.35% of range	Current 0.15% of reading + 0.35% of range 0.15% of reading + 0.35% of range 0.20% of reading + 0.35% of range	Power 0.35% of reading + 0.50% of range 0.25% of reading + 0.50% of range 0.40% of reading + 0.50% of reading		
* 50Hz/60Hz Compliant IEC61000-4-7						
D/A OUTPUT (OPTIONS)						
Item	Specifications					
Output Voltage	±5 V FS (approach ±7.5 \)	/ maximum) against e	each rated value.			
Number Of Output Channels Output Items		. P. VA. VAR. PE DEG	VHZ, IHZ, Vnk, Ink	WP, WP±, a, a+, Off		
Accuracy	Set for each channel: V, I, P, VA, VAR, PF, DEG, VHZ, IHZ, Vpk, Ipk, WP, WP±, q, q±, Off ±(accuracy of each measurement item + 0.2% of FS) (FS = 5 V)					
D/A Conversion Resolution	16 bits					
Minimum Load	100 kΩ					
Update Interval	Same as the data update In the case of Auto Upda	interval. Ite Rate, update interv	al is equal to signal in	nterval. More than 100ms		
Temperature Coefficient	In the case of Auto Update Rate, update interval is equal to signal interval. More than 100ms. ±0.05%/°C of FS					
REMOTE CONTROL INPUT/OUTPUT	'					
Item	Specifications					
Remote Control Input Signal	EXT HOLD, EXT TRIG, EXT START, EXT STOP, EXT RESET					
Remote Control Output Signal	INTEG BUSY					
I/O Level I/O Logic Format	TTL Negative logic, Falling edge					
1/0 Logic Format	INEGATIVE TOGIC, Failing ec	ige				

SPECIFICATIONS DIGITAL IO SIGNAL (OPTIONS) Specifications I/O Control Output Signal OUT1, OUT2, OUT3, OUT4 I/O Level TTI I/O Sink Current

- * Q (VAR), S (VA), λ (PF) and Φ (DEG) are originated from the measured values including voltage, current and active power which go through computation process. In respect to distorted signal input, accordingly, the value acquired from other instruments, which employ different methods, may differ from that acquired from GPM-8310 unit.
- * "Zero" will be shown for S or Q and "--" will be displayed for λ and Φ when either current or voltage is less than 0.5% of the rated range (less than or equivalent to 1% when crest factor is set 6).

GENERAL



The below are the basic conditions required to operate the GPM-8310 within specifications:

- 1-year Calibration: Yearly
- Operating Environment: 18~28 °C (64.4~82.4 °F)
- Humidity: <80%RH,

Max 100mA (per/ch)

- Accuracy: ± (% of reading + % of range)
- The specifications apply when it warmed up for at least 30 minutes and operates in the slow rate.
- The power supply cable must be grounded to ensure accuracy.
- · Input voltage and current must be standard sine wave.
- The power factor must be 1.
- The crest factor must be 3.
- The common-mode voltage must be zero.

Specification Condition	n
Operation Condition	

Temperature: 23°C±5°C

Humidity: <80%RH(non-condensing)

Temperature 0°C ~ 40°C, • 30 ~ 40°C, Relative Humidity < 70%RH (non-condensing) • >40°C, Relative Humidity < 50%RH (non-condensing) Indoor use only

Altitude: < 2000 meters Pollution degree 2

Temperature -40°C ~ 70°C **Storage Condition** Humidity: < 90%RH (non-condensing)

Power Source AC 100-240V, 50-60Hz; Consumption Max. 30VA 268(W) x 107(H) x 379(D) mm (w/t bumpers) Dimensions

Weight Approx. 2.9kg

PM-8213CD1BH Specifications subject to change without notice.

DA4 Interface (including cable, GTL-214)

ORDERING INFORMATION

GPM-8310 Digital Power Meter

with RS-232C/USB device & host/LAN/GPIB

GPM-8310 with DA4 Digital Power Meter

with RS-232C/USB device & host/LAN/GPIB and opt. DA4

ACCESSORIES

Safety Instruction Sheet x 1, Power cord x 1 Test lead GTL-209 x 1, Test lead GTL-212 x 1

CD x 1 (including complete user manual and USB driver) DA4 cable GTL-214 (available for GPM-8310 with DA4 only)

OPTION ACCESSORIES GPM-001 Test Fixture (including GTL-210 x 2, GTL-213 x 1) GPM-001(EU) Test Fixture (including GTL-210 x 2, GTL-213 x 1) GTI -209 Test Lead, Banana to Bare-wire, Approx. 1000mm

Note: Optional DA4 interface must be installed in factory.

OPTION Opt.01

GTL-210 Test Lead, Banana to Banana, Approx. 1000mm GTL-212 Test Lead, O-Type to Bare-wire, Approx. 1000mm GTI-213 Test Lead, O-Type to Banana, Approx. 1000mm GTI-214 DA4 Cable, Approx. 1000mm

GTL-232 RS-232C cable, 9-pin Female to 9-pin, null modem for computer, Approx. 2000mm USB Cable, A-B type, Approx. 1200mm GTL-246

GTL-248 GPIB Cable, Approx. 2000mm GRA-422 Rack Mount Kit, 19" 2U size

