

 Test Equipment
 5 Commonwealth Ave

 Woburn, MA 01801
 Phone 781-665-1400

 1-800-517-8431
 Toll Free 1-800-517-8431

 Visituas at www.festEquipmentDepot.com

pH Electrodes

<u>EN-1</u>

Model	Item NO	Description
ST310	83033965	3-in-1 pH electrode
ST210	83033966	2-in-1 pH electrode
ST320	83033967	3-in-1 gel pH electrode
ST230	83033968	Muddy sample pH electrode
STPURE	83033969	Pure water pH electrode

Specifications:

Connection:	BNC, Cinch		
Cable length:	1.0m		
Shaft length:	120mm		
Shaft diameter:	12mm		
pH range:	0-14pH		
Zero potential:	7.00±0.25pH		
Slope:	≥ 97%		
Impedance:	≤ 250MΩ		
Shaft material:			
	10,ST310,ST320)		
Glass (ST23	0, STPURE)		
Reference Electrolyte:			
3M KCI Gel (ST320) or Solution		
Temperature range:			
0-80℃ (ST2´	10,ST310,ST320)		
0-100℃ (ST2	230, STPURE)		
Temperature probe: NTC 30kΩ			
Reference Diaphragm:			
Ceramic (ST	210, ST310)		
Fiber (ST320))		
Sleeve or gro	ound glass		
(ST230, STP	'URE)		

<u>EN-2</u>

pH Electrodes

OHAUS offer different pH electrode for different application. Make sure you have chosen the best fit pH electrode for your samples; and do calibration first for the electrode.

- 3-in-1 electrodes(ST310 and ST320) have the temperature probe in it.
- Gel electrode(ST320) doesn't need to refill the reference electrolyte.
- 2-in-1 pH electrodes(ST210, ST230, STPURE) need to work with STTEMP30 for 0.01 pH accuracy.

Before using electrode in calibration or measurement, we should note:

1. When unpacking, check the electrode for any mechanical damage to the membrane, diaphragm or body.

2. Screw off watering cap first, then the lid, rinse electrode with pure water. Wipe electrode with a tissue. Do not rub the glass membrane, or the response time might increase.

3. Eliminate any air bubbles inside the membrane. Make the electrode inside pressure equilibrium with the atmosphere by opening the electrode refilling aperture when measurement.

pH Electrodes

FAQs:

1. What's the meaning of electrode condition icon?

After calibration, the pH meter will judge the electrode slope and offset, then display the judgment in the screen:

Slope > 95%

Electrode condition is very good

Slope : 90%-95%

Electrode condition is acceptable

⁽²⁾ slope : 85%-90%

Electrode condition is not good, need recalibration or replace

2.How do you properly store the pH electrode after every use?

A: Rinse the electrode with pure water, put the electrode into the watering cap. If the solution in the cap is polluted or dry, clean it and fill it with 3M KCl solution.

3.What's the life of the pH electrodes?

A: Measured samples, storage conditions and maintenance methods all could influence the life of pH electrodes; normally a pH electrode expires after 12 months.