# Test Equipment Depot - 800.517.8431 - TestEquipmentDepot.com

#### Safety terms in this manual

CAUTION. These statements identify conditions or practices that could result in damage to the equipment or other property.

#### Safety symbols and terms on the product

DANGER indicates an injury hazard immediately accessible as you read the marking.

WARNING indicates an injury hazard not immediately accessible as you read the marking.

CAUTION indicates a hazard to property including the product.



# CAUTION Refer to Manual

# **Compliance information**

The WFM200BA battery, including all battery options, meets the following requirements unless otherwise noted. Compliance demonstrated using high quality, shielded interface cables.

### **European Union**

#### EMC Directive 2004/108/EC.

- EN 55103-1:2009 Environment E2 – Product family standard for audio, video, audio-visual. Industrial and light commercial. Part 1: Emission.
  - EN 55022:2006 Class B Radiated and conducted emissions
  - EN 55103-1:2009 Annex A Radiated magnetic field emissions
- EN 55103-1:2009 Environment E2 Product Family Standard for Audio, Video, Audio-visual. Industrial and Light Commercial. Part 2: Immunity.
  - EN 55103-2:2009 Annex A Radiated Magnetic Field Immunity
  - IEC 61000-4-11:2004 Power Line Voltage Dip and Interruption Immunity
  - IEC 61000-4-2:2000 Electrostatic Discharge Immunity
  - IEC 61000-4-3:2007 RF Electromagnetic Field Immunity
  - IEC 61000-4-4:2004 Electrical Fast Transient / Burst Immunity
  - IEC 61000-4-5:2005 Power Line Surge Immunity
  - IEC 61000-4-6:2006 RF Conducted Immunity -
- EN 61000-3-2:2006 AC Power Line Harmonic Emissions
- EN61000-3-3:1995 - Voltage Changes, Fluctuations, and Flicker

### European contact. Tektronix UK, Ltd.

Western Peninsula Western Road Bracknell, RG12 1RF United Kingdom

#### Australia / New Zealand

#### Australia Radiocommunications Act 1992.

EN 55022:2006 Class B - Radiated and conducted emissions, in accordance with EN 55103-1:2009

#### Australia / New Zealand contact.

- Baker & McKenzie Level 27, AMP Centre 50 Bridge Street
- Sydney NSW 2000, Australia

#### Battery recycling

This WFM200BA Lithium-Ion rechargeable battery pack must be recycled or disposed of properly.

Lithium-Ion batteries are subject to disposal and recycling regulations that vary by country and region. Always check

#### Lithium-lon battery transportation

The capacity of the WFM200BA lithium ion battery pack is under  $\hat{1}00$  Wh. The battery meets the applicable requirements of UN Manual of Tests and Criteria Part III Section 38.3. As shipped from Tektronix, the battery quantity is under the limit for shipment according to Section II of the relevant Packing Instructions from the IATA Dangerous Goods Regulations. Consult your air carrier for applicability and determination of any special lithium battery transportation requirements.


- Always check all applicable local, national, and international regulations before transporting a Lithium-Ion battery.
- Transporting an end-of-life, damaged, or recalled battery may, in certain cases, be specifically limited or prohibited.
- The battery pack must be adequately protected against short-circuit or damage during transport.

#### Description

The WFM200BA rechargeable battery pack is intended for use with specifically-designed Tektronix products. The battery consists of an array of LiIon cells and an internal battery pack controller device, including internal fault protection.

Refer to your instrument's documentation to ensure this is the proper battery for your instrument.

When the battery is not installed in an instrument, you can check the battery charge level by pressing the Check button on the back of the battery pack. LEDs illuminate to indicate the amount of charge in increments of approximately 20%.



**NOTE.** For optimum performance, charge the battery pack completely before using it for the first time or after prolonged storage.

#### Battery maintenance guidelines

The WFM200BA Lithium-Ion rechargeable battery pack requires routine maintenance and care in its use and handling. Be sure to use the following guidelines to safely use the WFM200BA Lithium-Ion batteries and achieve the maximum battery life span.

#### Battery pack considerations

Do not leave a battery unused for an extended period of time, either in the product or in storage. When a battery has been unused for six months, check the charge status, and charge or dispose of the battery as appropriate. See Charging the battery pack and Battery recycling.

The typical estimated life of a Lithium-Ion battery is about two to three years, or 300 to 500 charge cycles, whichever occurs first. One charge cycle is a period of use from fully charged, to fully discharged, and fully recharged again. You should not expect a two to three year life expectancy for batteries that do not run through complete charge cycles.

A rechargeable Lithium-Ion battery has a limited life and will gradually lose its capacity to hold a charge. This loss of capacity (aging) is irreversible. As the battery loses capacity, the length of time it will power the product (run time) decreases.

A Lithium-Ion battery continues to slowly discharge (self-discharge) when not in use or while in storage. You need to routinely check the charge status of the battery.

Use the WFM200BA battery pack only in approved Tektronix instruments.

# **WFM200BA Rechargeable Battery Pack**

## Instructions



075-1041-04

# General safety summary

precautions to avoid injury and prevent damage to this product or any products connected to it. Carefully read all instructions. Retain these instructions for future reference.

Comply with local and national safety codes.

For correct and safe operation of the product, it is essential that you follow generally accepted safety procedures in addition to the safety precautions specified in this manual.

The product is designed to be used by trained personnel only.

This product is not intended for detection of hazardous voltages.

#### To avoid fire or personal injury

Do not operate with suspected failures. If you suspect there is damage to the battery, charger, or power supply, have it inspected by qualified service personnel.

Disable the product if it is damaged. Do not use the product if it is damaged or operates incorrectly. If in doubt about safety of the product, turn it off and disconnect the power cord. Clearly mark the product to prevent its further operation.

Examine the exterior of the product before you use it. Look for cracks or missing pieces.

Tektronix

Use the product only as specified. Review the following safety

Recharge battery pack properly. Recharge the battery pack only for the recommended charge cycle at the recommended temperature.

**Replace battery pack properly.** Replace the battery pack only with the specified type and rating.

Use proper battery charger. Use only the WFM200BC battery charger or approved Tektronix instrument to charge the WFM200BA battery pack.

Do not operate in wet/damp conditions. Be aware that condensation may occur if a unit is moved from a cold to a warm environment.

Do not operate in an explosive atmosphere.

Keep product surfaces clean and dry.

Do not attempt to service this product. There are no serviceable components in the WFM200BA battery.

and follow your applicable regulations before disposing of any battery. Contact Rechargeable Battery Recycling Corporation for U.S.A. and Canada, or your local battery recycling organization.

- Many countries prohibit the disposal of waste electronic equipment in standard waste receptacles.
- Place only discharged batteries in a battery collection container. Use electrical tape or other approved covering over the battery connection points to prevent short circuits.



This symbol indicates that this product complies with the applicable European Union requirements according to Directives 2012/19/EU and 2006/66/EC on waste electrical and electronic equipment (WEEE) and batteries. For information about recycling options, check the Tektronix Web site

#### Battery pack maintenance

- Observe and note the run time that a new fully-charged battery provides for powering your product. You can use this new battery run time as a basis to compare run times for older batteries. The run time of your battery will vary depending on the product's configuration and the applications that you run.
- The battery pack gives the best performance when operated at normal room temperature,  $20 \text{ °C} \pm 5 \text{ °C} (68 \text{ °F} \pm 9 \text{ °F})$ .
- Routinely check the charge status of the battery.
- Carefully monitor a battery that is approaching the end of its estimated life.
- Consider replacing the battery with a new one if the battery run time drops below about 80% of the original run time, or the battery charge time increases significantly.
- Follow the storage requirements if you store or do not use a battery for an extended period. If you do not follow the storage requirements, and the battery will not power the instrument when installed, consider the battery to be damaged. Do not connect power to the instrument during this time. Remove the battery pack from the instrument. Do not attempt to recharge or reuse the battery. Replace it with a new battery.

#### Battery pack charging

The battery pack charges automatically when installed in an approved Tektronix instrument that is connected to an external power supply. You can also charge the battery pack with the Tektronix WFM200BC External Battery Charger. Remove the battery pack from the WFM200BC charger when charging is complete.

A Lithium-Ion battery pack will self-discharge during non-use. To achieve the longest operating time, charge your battery pack before use. If you plan to store battery packs, read the instructions in the *Battery pack storage* section of this document.

Configuration	Typical charging time
One fully discharged battery pack charging in the instrument	5.5 hours, typical. Time varies per instrument type and the instrument's power usage while charging.
One or two fully discharged battery packs charging in the WFM200BC external battery charger	5.5 hours

**CAUTION.** To avoid damage to the battery pack, use only the WFM200BC charger or an approved instrument to charge the battery pack. Do not connect any other voltage source to the battery pack.

**NOTE.** For optimum performance, charge the battery pack completely before using it for the first time or after prolonged storage.

To discharge the battery pack completely, continue to run the instrument on the battery pack until automatic shutdown occurs. After the instrument shuts down, either recharge the battery pack promptly or remove it from the instrument until you are ready to recharge the battery pack.

**NOTE.** To prolong the life of the battery pack and to prevent shutdown, do not operate or charge the battery pack at high temperatures. For best results, allow the battery pack to cool to room temperature before using or charging the battery pack.

Do not leave a battery pack discharged for extended periods. See the Storing battery packs section for information on how to properly store a battery pack.

The battery pack stops charging under the following conditions:

- The battery pack is fully charged.
- The internal temperature of the battery pack exceeds the safe charging threshold.
- A fault condition occurs.

- Do not short the external contacts on the battery.
- Do not dispose of the battery in fire or water.
- Do not expose the battery to temperatures above +60 °C (+140 °F).
- Seek medical advice if a battery or part of it has been swallowed.
- Do not put battery packs near heat or fire. Do not put in sunlight.
- Keep the battery pack clean and dry. Clean dirty connectors with a dry, clean cloth.
- Batteries contain hazardous chemicals that can cause burns or explode. If exposure to chemicals occurs, clean with soap and water and get medical aid. Repair the product before use if the battery leaks.
- Do not attempt to open, modify, reform or repair a battery pack that appears to be malfunctioning, or which has been physically damaged.
- Keep the battery away from children and animals.
- Avoid exposing the battery to excessive shock or vibration.
- Do not use a damaged battery or charger.
- If a battery pack has leaking fluids, do not touch any fluids. Dispose of a leaking battery pack. See the *Battery recycling* section for information on disposal and recycling.
- In case of eye contact with fluid, do not rub eyes. Immediately flush eyes thoroughly with water for at least 15 minutes, lifting upper and lower lids, until no evidence of the fluid remains. Seek medical attention.

#### Battery pack storage

- Store battery packs in a low-humidity environment free of corrosive gases. Storing battery packs in high-humidity environments, or outside the temperature range, can cause oxidation on the metallic parts, increased leakage, significant aging and premature failure.
- Do not store battery packs near heat or fire. Do not store in the sunlight.
- Do not keep battery packs in a place where the terminals can be shorted by metal objects (for example, coins, paper clips, or pens).
- Do not remove a battery pack from its original packaging until required for use.
- When possible remove the battery pack from the equipment when not in use.
- Charge the battery to approximately 50% of capacity at least once every six months.
- Remove the battery and store it separately from the instrument. Refer to the documentation for your instrument for instructions on removing and installing the battery.
- For optimal battery life, store the battery at temperatures between +5 °C and +25 °C (+41 °F and +77 °F). Higher temperatures reduce the battery storage life.

#### Battery pack replacement

Replace the Lithium-Ion battery pack only with a WFM200BA battery pack.

See the *Battery recycling* section for information on disposal and recycling.

# Specifications

opeeneere	
Characteristic	Description
Capacity, typical	6.14 Ah
Output, nominal	14.4 VDC
Charging temperature	10 °C to +40 °C (+50 °F to +104 °F); varies per the instrument's heat dissipation characteristics, the actual limit may be lower
Operating temperature	Installed in the instrument: 0 °C to +40 °C (+32 °F to +104 °F)
Storage temperature	-20 °C to +60 °C (-4 °F to +140 °F)
Weight	Battery: 453.6 g (16.0 oz)
	Battery (including packaging): 570 g (20.1 oz)

The battery pack shuts down and disconnects itself from the instrument under the following conditions:

- The battery pack is fully discharged.
- The internal temperature of the battery pack exceeds the safe discharging threshold.
- A fault condition occurs.

# Battery pack handling

Do not disassemble, crush, drop, or puncture the battery.